


LB106599
INSTRUMENT ID CV1

7471B



Linear

A= 0.0000e+000

B= 1.6904e-004

C= -1.7750e-002

Rho= 0.9999333

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	236	0.000	236				
0.2	0.200	0.201	0.001	1296	0.0 %	1296				
2.5	2.500	2.503	0.003	14912	0.0 %	14912				
5.0	5.000	4.920	-0.080	29210	0.0 %	29210				
7.5	7.500	7.562	0.062	44843	0.0 %	44843				
10.0	10.000	9.992	-0.008	59214	0.0 %	59214				

LB106599 INSTRUMENT ID CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Dec 2019 07:54:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO.0	236	-	0.0000	7471B	PPB	05 Dec 2019 07:56:58	S	Std
0.2 - 1	SO.2	1296	-	0.2000	7471B	PPB	05 Dec 2019 07:59:16	S	Std
2.5 - 1	S2.5	14912	-	2.5000	7471B	PPB	05 Dec 2019 08:01:33	S	Std
5.0 - 1	SS.0	29210	-	5.0000	7471B	PPB	05 Dec 2019 08:03:50	S	Std
7.5 - 1	S7.5	44843	-	7.5000	7471B	PPB	05 Dec 2019 08:06:07	S	Std
10.0 - 1	S10.0	59214	-	10.0000	7471B	PPB	05 Dec 2019 08:08:23	S	Std
ICV17 - 1	ICV17	23355	3.9301	-	7471B	PPB	05 Dec 2019 08:14:33	U	SMPL
ICB17 - 1	ICB17	26	-0.0134	-	7471B	PPB	05 Dec 2019 08:16:48	U	SMPL
CCV69 - 1	CCV69	28860	4.8606	-	7471B	PPB	05 Dec 2019 08:19:04	U	SMPL
CCB69 - 1	CCB69	44	-0.0103	-	7471B	PPB	05 Dec 2019 08:21:19	U	SMPL
CRA - 1	CRA	1354	0.2111	-	7471B	PPB	05 Dec 2019 08:23:34	U	SMPL
HighStd - 1	HighStd	57896	9.7687	-	7471B	PPB	05 Dec 2019 08:25:50	U	SMPL
ChkStd - 1	ChkStd	40876	6.8918	-	7471B	PPB	05 Dec 2019 08:28:05	U	SMPL
PB125246BL - 1	PBS	26	-0.0134	-	7471B	PPB	05 Dec 2019 08:30:23	U	SMPL
PB125246BS - 1	LCSS	23136	3.8931	-	7471B	PPB	05 Dec 2019 08:32:40	U	SMPL
K6100-01 - 1	TH-1	4892	0.8092	-	7471B	PPB	05 Dec 2019 08:34:56	U	SMPL
K6100-01DUP - 1	TH-1DUP	4985	0.8249	-	7471B	PPB	05 Dec 2019 08:41:23	U	SMPL
K6100-01MS - 1	TH-1MS	30631	5.1600	-	7471B	PPB	05 Dec 2019 08:43:40	U	SMPL
K6100-01MSD - 1	TH-1MSD	30760	5.1818	-	7471B	PPB	05 Dec 2019 08:45:57	U	SMPL
K6100-03 - 1	TH-9	2040	0.3271	-	7471B	PPB	05 Dec 2019 08:48:13	U	SMPL
CCV70 - 1	CCV70	28873	4.8628	-	7471B	PPB	05 Dec 2019 08:50:28	U	SMPL
CCB70 - 1	CCB70	2	-0.0174	-	7471B	PPB	05 Dec 2019 08:52:44	U	SMPL
K6112-01 - 1	OK-02-120419	1450	0.2274	-	7471B	PPB	05 Dec 2019 08:54:59	U	SMPL
K6119-01 - 1	SB-3-(0-2)	1408	0.2203	-	7471B	PPB	05 Dec 2019 08:57:14	U	SMPL
K6119-02 - 1	SB-3-(8-10)	250	0.0245	-	7471B	PPB	05 Dec 2019 08:59:30	U	SMPL
K6119-03 - 1	SB-6-(1-3)	11761	1.9703	-	7471B	PPB	05 Dec 2019 09:01:46	U	SMPL
K6119-04 - 1	SB-6-(7-9)	572	0.0789	-	7471B	PPB	05 Dec 2019 09:04:02	U	SMPL
K6119-05 - 1	SB-4-(0-2)	550	0.0752	-	7471B	PPB	05 Dec 2019 09:06:19	U	SMPL
K6119-06 - 1	SB-4-(8-10)	166	0.0103	-	7471B	PPB	05 Dec 2019 09:08:36	U	SMPL
K6119-07 - 1	SB-5-(0-2)	488	0.0647	-	7471B	PPB	05 Dec 2019 09:10:53	U	SMPL
K6119-08 - 1	SB-5-(8-10)	326	0.0374	-	7471B	PPB	05 Dec 2019 09:13:10	U	SMPL
K6119-09 - 1	SB-2-(0-2)	5185	0.8587	-	7471B	PPB	05 Dec 2019 09:15:25	U	SMPL
CCV71 - 1	CCV71	28530	4.8048	-	7471B	PPB	05 Dec 2019 09:51:09	U	SMPL
CCB71 - 1	CCB71	54	-0.0086	-	7471B	PPB	05 Dec 2019 09:53:24	U	SMPL
K6119-10 - 1	SB-2-(8-10)	226	0.0205	-	7471B	PPB	05 Dec 2019 09:55:40	U	SMPL
K6119-11 - 1	SB-1-(0-2)	2599	0.4216	-	7471B	PPB	05 Dec 2019 09:57:56	U	SMPL
K6119-12 - 1	SB-1-(8-10)	-18	-0.0208	-	7471B	PPB	05 Dec 2019 10:00:11	U	SMPL
K6120-01 - 1	TH-7	893	0.1332	-	7471B	PPB	05 Dec 2019 10:02:26	U	SMPL
K6120-03 - 1	TH-6	1423	0.2228	-	7471B	PPB	05 Dec 2019 10:04:42	U	SMPL
K6120-05 - 1	TH-4	850	0.1259	-	7471B	PPB	05 Dec 2019 10:06:58	U	SMPL
K6120-07 - 1	TH-5	47	-0.0098	-	7471B	PPB	05 Dec 2019 10:09:15	U	SMPL
K6100-01LX5 - 1		1084	0.1655	-	7471B	PPB	05 Dec 2019 10:11:29	U	SMPL
CCV72 - 1	CCV72	27345	4.6045	-	7471B	PPB	05 Dec 2019 10:13:46	U	SMPL
CCB72 - 1	CCB72	183	0.0132	-	7471B	PPB	05 Dec 2019 10:16:02	U	SMPL