



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106606

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/05/2019	11:35
2	CAL2	1	Water	NA	NA	20.1	7.01	12/05/2019	11:36
3	CAL3	1	Water	NA	NA	20.3	10.02	12/05/2019	11:37
4	ICV	1	Water	NA	NA	20.1	7.01	12/05/2019	11:39
5	CCV1	1	Water	NA	NA	20.2	2.01	12/05/2019	11:40
6	K6163-01	1	Water	NA	NA	18.9	4.32	12/05/2019	11:42
7	K6173-01	1	Water	NA	NA	20.4	6.03	12/05/2019	11:43
8	K6173-01DUP	1	Water	NA	NA	20.5	6.04	12/05/2019	11:44
9	CCV2	1	Water	NA	NA	20.1	12.02	12/05/2019	11:45

20106606

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6163 WorkList ID : 133902 Department : Wet-Chemistry Date : 12-05-2019 10:56:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2019	Water	K6163-01 - ☒	pH	Cool 4 deg C	STER05	H12	BOTTLE-1	12/04/2019	SM 4500-H B
12/04/2019	Water	K6173-01 - ☒	pH	Cool 4 deg C	STER01	H12	POOL-WATER	12/04/2019	SM 4500-H B

Date/Time 12.05.19 11:30
Received by: JP
Relinquished by: CP

Date/Time 12.05.19 12:30
Received by: CP
Relinquished by: JP