



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106608

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	12/05/2019	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	12/05/2019	13:36
3	CAL3	1	Water	NA	NA	20.1	10.02	12/05/2019	13:37
4	ICV	1	Water	NA	NA	20.1	7.01	12/05/2019	13:38
5	CCV1	1	Water	NA	NA	20.1	2.01	12/05/2019	13:39
6	K6174-05	1	Water	NA	NA	12.6	11.07	12/05/2019	13:40
7	K6174-05DUP	1	Water	NA	NA	12.7	11.08	12/05/2019	13:41
8	CCV2	1	Water	NA	NA	20.1	12.02	12/05/2019	13:42

28 106608

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K6174 WorkList ID : 133915 Department : Wet-Chemistry Date : 12-05-2019 13:25:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2019	Water	K6174-05	Corrosivity	Cool 4 deg C	PSEG01	N21	WATER-COMP	12/05/2019	9040C

Date/Time 12-05-19 13:30
Received by: Jv CP
Relinquished by: _____

Date/Time 12-05-19 14:30
Received by: CP
Relinquished by: Jv