



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106613

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/05/2019	10:25
2	K6134-01		1	9	100.00	>212.0	>212.0	12/05/2019	11:50
3	K6134-01DUP		1	9	100.00	>212.0	>212.0	12/05/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 106613

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP 12519 WorkList ID : 133896 Department : Wet-Chemistry Date : 12-05-2019 09:53:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2019	Water	K6134-01	H Flash Point	Cool 4 deg C	PSEG01	N13	30573	12/04/2019	1010A

Date/Time 12/05/19 09:55
Received by: JS
Relinquished by: JP

Date/Time 12/05/19 12:25
Received by: JP
Relinquished by: KS