



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/10/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 16:19
In Date: 12/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 12/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106624

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6156-01	P001-CF41-3036-01	1	1.14	9.55	9.31	97.1	
K6156-02	P001-CF42-0612-01	2	1.14	9.93	9.16	91.2	
K6156-03	P001-CF43-0612-01	3	1.13	9.64	8.91	91.4	
K6156-04	P001-CF44-0612-01	4	1.14	9.62	8.96	92.2	
K6156-05	P001-CF45-0612-01	5	1.14	9.55	8.88	92.0	
K6156-06	P001-CF46-0612-01	6	1.15	9.73	9.00	91.5	
K6156-07	P001-CF47-0612-01	7	1.15	9.68	9.15	93.8	
K6156-08	P001-CF48-0612-01	8	1.15	9.72	9.06	92.3	
K6156-09	P001-CF49-0612-01	9	1.16	9.67	8.92	91.2	
K6156-10	P001-CF49-0612-01MS	10	1.16	9.67	8.92	91.2	
K6156-11	P001-CF49-0612-01MSD	11	1.16	9.67	8.92	91.2	
K6156-12	P001-CF49-0612-02	12	1.15	9.6	8.85	91.1	
K6156-13	P001-CF50-0006-01	13	1.14	9.7	8.63	87.5	
K6156-14	P001-CG41-1824-01	14	1.16	9.83	9.14	92.0	
K6156-15	P001-CG42-1218-01	15	1.15	9.76	8.94	90.5	
K6156-16	P001-CG43-1824-01	16	1.13	9.94	9.26	92.3	
K6156-17	P001-CG44-1218-01	17	1.14	9.98	9.42	93.7	
K6156-18	P001-CG45-0612-01	18	1.15	9.9	9.2	92.0	
K6156-19	P001-CG46-1218-01	19	1.16	9.64	9.1	93.6	
K6156-20	P001-CG47-1218-01	20	1.14	9.54	8.86	91.9	
K6156-21	P001-CG48-1218-01	21	1.15	9.55	9.00	93.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$