



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/10/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 16:38
In Date: 12/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 12/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106626

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K6158-01	P001-CG49-0006-01	1	1.14	9.82	9.14	92.2	
K6158-02	P001-CG50-0006-01	2	1.14	9.9	8.85	88.0	
K6158-03	P001-CH45-0612-01	3	1.15	9.73	9.19	93.7	
K6158-04	P001-CH46-0612-01	4	1.15	9.82	9.26	93.5	
K6158-05	P001-CH47-0612-01	5	1.16	9.79	9.12	92.2	
K6158-06	P001-CH48-0612-01	6	1.14	9.76	9.09	92.2	
K6158-07	P001-CH49-0612-01	7	1.12	9.52	8.84	91.9	
K6158-08	P001-CH50-0006-01	8	1.12	9.93	9.03	89.8	
K6158-09	P001-CI45-1218-01	9	1.14	9.73	9.23	94.2	
K6158-10	P001-CI45-1218-01D	10	1.14	9.73	9.23	94.2	
K6158-11	P001-CI45-1218-01S	11	1.14	9.73	9.23	94.2	
K6158-12	P001-CI45-1218-02	12	1.14	9.55	9.07	94.3	
K6158-13	P001-CI46-1824-01	13	1.13	9.76	9.15	92.9	
K6158-14	P001-CI47-0612-01	14	1.14	9.8	9.06	91.5	
K6158-15	P001-CI48-0612-01	15	1.13	9.9	8.92	88.8	
K6158-16	P001-CI49-1824-01	16	1.14	9.7	8.96	91.4	
K6158-17	P001-CI50-0006-01	17	1.14	9.77	8.98	90.8	
K6158-18	P001-CJ45-0612-01	18	1.12	9.71	9.09	92.8	
K6158-19	P001-CJ46-0612-01	19	1.13	9.74	9.07	92.2	
K6158-20	P001-CJ47-0612-01	20	1.14	9.67	8.99	92.0	
K6158-21	P001-CJ48-1218-01	21	1.13	9.82	9.06	91.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$