



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106635

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/06/2019	11:00
2	CAL2	1	Water	NA	NA	20.1	7.00	12/06/2019	11:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/06/2019	11:02
4	ICV	1	Water	NA	NA	20.1	7.01	12/06/2019	11:04
5	CCV1	1	Water	NA	NA	20.1	2.01	12/06/2019	11:05
6	K6189-01	1	Water	NA	NA	18.1	4.96	12/06/2019	11:06
7	K6189-01DUP	1	Water	NA	NA	18.2	4.97	12/06/2019	11:07
8	CCV2	1	Water	NA	NA	20.1	12.02	12/06/2019	11:10

LB 106635

WORKLIST(Hardcopy Internal Chain)

Date : 12-06-2019 10:47:28

Department : Wet-Chemistry

WorkList ID : 133960

WorkList Name : PH K6189

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2019	Water	K6189-01 - X	pH	Cool 4 deg C	STER14	H13	JS-10000-SYSTEM	12/04/2019	SM 4500-H B

Date/Time 12-06-19 10:50
Received by: SN
Relinquished by: ST

Date/Time 12-06-19 11:50
Received by: ST
Relinquished by: ST