



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/9/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 12/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 07:40
Out Date: 12/09/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106645

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6149-01	DBBY9	1	1.00	2.00	2.00	100.0	P.E.SAMPLE
K6149-02	DBBL1	2	1.15	9.89	7.81	76.2	
K6149-03	DBBL1MS	3	1.15	9.89	7.81	76.2	
K6149-04	DBBL1MSD	4	1.15	9.89	7.81	76.2	
K6149-05	DBBL3	5	1.16	9.77	8.24	82.2	
K6149-06	DBBL4	6	1.18	9.93	8.41	82.6	
K6149-07	DBBR8	7	1.13	9.77	8.72	87.8	
K6149-08	DBBS0	8	1.15	9.96	8.95	88.5	
K6149-09	DBBS2	9	1.14	9.58	8.68	89.3	T.B.
K6149-10	DBBS3	10	1.16	9.91	7.55	73.0	
K6149-11	DBBS4	11	1.15	9.66	8.37	84.8	
K6149-12	DBBS5	12	1.12	9.92	8.51	84.0	
K6149-13	DBBS7	13	1.14	9.63	8.2	83.2	
K6149-14	DBBS9	14	1.15	9.92	8.49	83.7	
K6149-15	DBBT0	15	1.18	9.97	8.65	85.0	
K6149-16	DBBT5	16	1.18	9.81	8.74	87.6	
K6149-17	DBBT6	17	1.13	9.51	8.26	85.1	
K6149-18	DBBT8	18	1.18	9.8	8.51	85.0	
K6149-19	DBBT9	19	1.14	9.65	8.2	83.0	
K6149-20	DBBW1	20	1.15	9.96	9.15	90.8	
K6149-21	DBBT3	21	1.17	9.98	8.65	84.9	
K6149-22	VHBLK01	22	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$