



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106650

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/06/2019	15:25
2	CAL2	1	Water	NA	NA	20.1	7.01	12/06/2019	15:26
3	CAL3	1	Water	NA	NA	20.1	10.03	12/06/2019	15:29
4	ICV	1	Water	NA	NA	20.1	6.99	12/06/2019	15:30
5	CCV1	1	Water	NA	NA	20.2	2.02	12/06/2019	15:32
6	K6194-02	1	Solid	20.03	20	21.3	10.16	12/06/2019	15:35
7	K6194-02DUP	1	Solid	20.04	20	21.4	10.17	12/06/2019	15:36
8	CCV2	1	Water	NA	NA	20.1	12.02	12/06/2019	15:39

HP106650

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K6194 S	WorkList ID :	134007	Department :	Wet-Chemistry	Date :	12-06-2019 15:10:04		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/06/2019	Solid	K6194-02	pH	Cool 4 deg C	AREC01	H61	WC-01	12/06/2019	9045D

Date/Time 12-06-19 15:15
Received by: SA
Relinquished by: SA

Date/Time 12-06-19 16:30
Received by: SA
Relinquished by: SA