



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106651

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/09/2019	12:15
2	CAL2	1	Water	NA	NA	20.2	7.00	12/09/2019	12:16
3	CAL3	1	Water	NA	NA	20.2	10.03	12/09/2019	12:17
4	ICV	1	Water	NA	NA	20.2	7.01	12/09/2019	12:19
5	CCV1	1	Water	NA	NA	20.1	2.02	12/09/2019	12:20
6	K6209-01	1	Water	NA	NA	19.7	4.61	12/09/2019	12:21
7	K6209-01DUP	1	Water	NA	NA	19.8	4.62	12/09/2019	12:22
8	K6211-01	1	Water	NA	NA	20.8	5.28	12/09/2019	12:24
9	CCV2	1	Water	NA	NA	20.1	12.02	12/09/2019	12:25

VB 166651

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6209 WorkList ID : 134006 Department : Wet-Chemistry Date : 12-09-2019 12:00:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2019	Water	K6209-01	pH	Cool 4 deg C	STER15	D11	GILROY-OMEGA-CELL	12/05/2019	SM 4500-H B
12/05/2019	Water	K6211-01	pH	Cool 4 deg C	STER04	D11	POOL-WATER	12/05/2019	SM 4500-H B

Date/Time 12-09-19 12:05
 Received by: J2
 Relinquished by: an

Date/Time 12-09-19 13:10
 Received by: an
 Relinquished by: J2