


LB106667
INSTRUMENT ID: CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	-80	0.000	-80				
0.2	0.200	0.159	-0.041	1010	0.0 %	1010				
2.5	2.500	2.523	0.023	13811	0.0 %	13811				
5.0	5.000	5.102	0.102	27770	0.0 %	27770				
7.5	7.500	7.568	0.068	41125	0.0 %	41125				
10.0	10.000	9.890	-0.110	53694	0.0 %	53694				

LB106667 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 10 Dec 2019 07:32:01				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	-80	-	0.0000	7471B	PPB	10 Dec 2019 07:40:23	S Std
0.2 - 1	30.2	1010	-	0.2000	7471B	PPB	10 Dec 2019 07:42:41	S Std
2.5 - 1	32.5	13811	-	2.5000	7471B	PPB	10 Dec 2019 07:44:58	S Std
5.0 - 1	35.0	27770	-	5.0000	7471B	PPB	10 Dec 2019 07:47:15	S Std
7.5 - 1	37.5	41125	-	7.5000	7471B	PPB	10 Dec 2019 07:49:32	S Std
10.0 - 1	38.0	53694	-	10.0000	7471B	PPB	10 Dec 2019 07:51:48	S Std
ICV20 - 1	ICV20	21901	4.0176	-	7471B	PPB	10 Dec 2019 07:56:43	U SMPL
ICB20 - 1	ICB20	0	-0.0275	-	7471B	PPB	10 Dec 2019 07:58:59	U SMPL
CCV79 - 1	CCV79	27184	4.9934	-	7471B	PPB	10 Dec 2019 08:01:14	U SMPL
CCB79 - 1	CCB79	20	-0.0238	-	7471B	PPB	10 Dec 2019 08:03:29	U SMPL
CRA - 1	CRA	1071	0.1703	-	7471B	PPB	10 Dec 2019 08:05:46	U SMPL
HighStd - 1	HighStd	53549	9.8630	-	7471B	PPB	10 Dec 2019 08:08:01	U SMPL
ChkStd - 1	ChkStd	37004	6.8072	-	7471B	PPB	10 Dec 2019 08:10:17	U SMPL
PB125337BL - 1	PBS	-23	-0.0317	-	7471B	PPB	10 Dec 2019 08:12:33	U SMPL
PB125337BS - 1	LCS	21996	4.0352	-	7471B	PPB	10 Dec 2019 08:14:50	U SMPL
K5955-01 - 1	CH-FLOOR	353874	HIGH	-	7471B	PPB	10 Dec 2019 08:17:06	X
K5955-03 - 1	CONTROL-HOUSE	19342	3.5450	-	7471B	PPB	10 Dec 2019 08:18:57	U SMPL
K5955-05 - 1	FOUNDATION-1	64	-0.0156	-	7471B	PPB	10 Dec 2019 08:21:16	U SMPL
K5955-07 - 1	FOUNDATION-2	86	-0.0116	-	7471B	PPB	10 Dec 2019 08:23:34	U SMPL
K6111-01 - 1	CL-01-120619	2979	0.5228	-	7471B	PPB	10 Dec 2019 08:25:50	U SMPL
CCV80 - 1	CCV80	27020	4.9631	-	7471B	PPB	10 Dec 2019 08:28:05	U SMPL
CCB80 - 1	CCB80	-14	-0.0301	-	7471B	PPB	10 Dec 2019 08:30:20	U SMPL
K6111-01DUP - 1	CL-01-120619DUP	3025	0.5312	-	7471B	PPB	10 Dec 2019 08:32:36	U SMPL
K6111-01MS - 1	CL-01-120619MS	21757	3.9910	-	7471B	PPB	10 Dec 2019 08:34:51	U SMPL
K6111-01MSD - 1	CL-01-120619MSD	21413	3.9275	-	7471B	PPB	10 Dec 2019 08:37:29	U SMPL
K6113-01 - 1	OR-03-120519	1358	0.2234	-	7471B	PPB	10 Dec 2019 08:39:44	U SMPL
K6117-01 - 1	TR-04-120619	3406	0.6016	-	7471B	PPB	10 Dec 2019 08:42:03	U SMPL
K6118-01 - 1	SU-02-120619	2181	0.3754	-	7471B	PPB	10 Dec 2019 08:44:19	U SMPL
K6133-01 - 1	ROLL-OFF-362	24249	4.4513	-	7471B	PPB	10 Dec 2019 08:46:36	U SMPL
K6135-02 - 1	COMP-34	5039	0.9032	-	7471B	PPB	10 Dec 2019 08:48:53	U SMPL
K6135-04 - 1	COMP-35	13999	2.5581	-	7471B	PPB	10 Dec 2019 08:51:11	U SMPL
K6135-06 - 1	COMP-36	11825	2.1566	-	7471B	PPB	10 Dec 2019 08:53:26	U SMPL
CCV81 - 1	CCV81	27310	5.0167	-	7471B	PPB	10 Dec 2019 08:55:42	U SMPL
CCB81 - 1	CCB81	-6	-0.0286	-	7471B	PPB	10 Dec 2019 08:57:57	U SMPL
K6141-04 - 1	CMJ-01-COMP	1248	0.2030	-	7471B	PPB	10 Dec 2019 09:00:13	U SMPL
K6145-01 - 1	TH-2	1255	0.2043	-	7471B	PPB	10 Dec 2019 09:02:29	U SMPL
K6145-03 - 1	TH-3	5464	0.9817	-	7471B	PPB	10 Dec 2019 09:04:44	U SMPL
K6145-05 - 1	TH-8	1206	0.1953	-	7471B	PPB	10 Dec 2019 09:07:00	U SMPL
K6197-01 - 1	WC-S	144	-0.0009	-	7471B	PPB	10 Dec 2019 09:09:16	U SMPL
PB125338BL - 1	PBS	23	-0.0232	-	7471B	PPB	10 Dec 2019 09:11:33	U SMPL
PB125338BS - 1	LCS	22263	4.0845	-	7471B	PPB	10 Dec 2019 09:13:50	U SMPL
K6147-01 - 1	P001-WC001	9375	1.7041	-	7471B	PPB	10 Dec 2019 09:16:07	U SMPL
K6147-01DUP - 1	P001-WC001DUP	9464	1.7205	-	7471B	PPB	10 Dec 2019 09:18:24	U SMPL
K6147-01MS - 1	P001-WC001MS	28533	5.2426	-	7471B	PPB	10 Dec 2019 09:20:40	U SMPL
CCV82 - 1	CCV82	27749	5.0978	-	7471B	PPB	10 Dec 2019 09:22:56	U SMPL
CCB82 - 1	CCB82	-3	-0.0280	-	7471B	PPB	10 Dec 2019 09:25:12	U SMPL
K6147-01MSD - 1	P001-WC001MSD	28791	5.2902	-	7471B	PPB	10 Dec 2019 09:27:28	U SMPL
K6147-02 - 1	P001-WC002	9763	1.7758	-	7471B	PPB	10 Dec 2019 09:29:44	U SMPL
K6147-03 - 1	P001-WC003	11951	2.1799	-	7471B	PPB	10 Dec 2019 09:31:59	U SMPL
K6147-04 - 1	P001-WC004	4996	0.8953	-	7471B	PPB	10 Dec 2019 09:34:15	U SMPL
K6147-05 - 1	P001-WC005	15305	2.7994	-	7471B	PPB	10 Dec 2019 09:36:31	U SMPL
K6147-06 - 1	P001-WC006	10561	1.9231	-	7471B	PPB	10 Dec 2019 09:38:47	U SMPL
K6147-07 - 1	P001-WC007	18624	3.4124	-	7471B	PPB	10 Dec 2019 09:41:04	U SMPL
K6147-08 - 1	P001-WC008	14224	2.5997	-	7471B	PPB	10 Dec 2019 09:43:21	U SMPL
K6147-09 - 1	P001-WC009	8062	1.4616	-	7471B	PPB	10 Dec 2019 09:45:38	U SMPL
K6147-10 - 1	P001-WC010	5402	0.9703	-	7471B	PPB	10 Dec 2019 09:47:54	U SMPL
CCV83 - 1	CCV83	28096	5.1619	-	7471B	PPB	10 Dec 2019 09:50:10	U SMPL
CCB83 - 1	CCB83	0	-0.0275	-	7471B	PPB	10 Dec 2019 09:52:26	U SMPL
K6147-11 - 1	P001-WC011	5571	1.0015	-	7471B	PPB	10 Dec 2019 09:54:42	U SMPL
K6162-01 - 1	TP1	37077	6.8206	-	7471B	PPB	10 Dec 2019 09:56:58	U SMPL
K6162-02 - 1	TP2	8629	1.5663	-	7471B	PPB	10 Dec 2019 09:59:15	U SMPL
K6174-01 - 1	SOLID-COMP	1356	0.2230	-	7471B	PPB	10 Dec 2019 10:01:31	U SMPL
K6188-01 - 1	SOIL-COMP-1	1540	0.2570	-	7471B	PPB	10 Dec 2019 10:03:48	U SMPL
K6188-05 - 1	SOIL-COMP-2	1207	0.1955	-	7471B	PPB	10 Dec 2019 10:06:04	U SMPL
K6194-02 - 1	WC-01	25854	4.7478	-	7471B	PPB	10 Dec 2019 10:08:21	U SMPL
K6111-01LX5 - 1		521	0.0688	-	7471B	PPB	10 Dec 2019 10:10:38	U SMPL
K6147-01LX5 - 1		1852	0.3146	-	7471B	PPB	10 Dec 2019 10:12:55	U SMPL
K5955-01DLX20 - 1		28684	5.2705	-	7471B	PPB	10 Dec 2019 10:19:13	U SMPL
CCV84 - 1	CCV84	28388	5.2158	-	7471B	PPB	10 Dec 2019 10:21:29	U SMPL
CCB84 - 1	CCB84	-12	-0.0297	-	7471B	PPB	10 Dec 2019 10:23:46	U SMPL