


LB106680
INSTRUMENT ID CV1

ISM02.4

Linear

μ Abs.:

40245



A= 0.0000e+000

B= 2.5043e-004 *Slope*

C= -3.0277e-002 *Intercept*

Rho= 0.9999270

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>1-D</i>
0.00	0.000	0.004	0.004	135	0.000	135					0
0.05	0.050										
0.20	0.200	0.197	-0.003	906	0.0 %	906					0
2.50	2.500	2.510	0.010	10142	0.0 %	10142					0
5.00	5.000	5.033	0.033	20219	0.0 %	20219					0
7.50	7.500	7.409	-0.091	29704	0.0 %	29704					-1
10.0	10.000	10.048	0.048	40245	0.0 %	40245					0

9
12/1/19

LB106680
INSTRUMENT ID CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Co	Method	Units	Date	Type
0	0.0	135	-	0	ISM02.4	PPB	12/11/2019 10:29	Std
0.2	0.2	906	-	0.2	ISM02.4	PPB	12/11/2019 10:31	Std
2.5	2.5	10142	-	2.5	ISM02.4	PPB	12/11/2019 10:34	Std
5	5.0	20219	-	5	ISM02.4	PPB	12/11/2019 10:36	Std
7.5	7.5	29704	-	7.5	ISM02.4	PPB	12/11/2019 10:38	Std
10	10.0	40245	-	10	ISM02.4	PPB	12/11/2019 10:41	Std
ICV	ICV	15426	3.8329 -		ISM02.4	PPB	12/11/2019 10:44	SMPL
ICB	ICB	0	-0.0303 -		ISM02.4	PPB	12/11/2019 10:47	SMPL
CCV001	CCV001	20263	5.0443 -		ISM02.4	PPB	12/11/2019 10:49	SMPL
CCB001	CCB001	4	-0.0293 -		ISM02.4	PPB	12/11/2019 10:51	SMPL
PB125382E	PBW382	31	-0.0225 -		ISM02.4	PPB	12/11/2019 10:54	SMPL
K6169-01	MC0BW4	23	-0.0245 -		ISM02.4	PPB	12/11/2019 10:56	SMPL
K6169-02	MC0BW4D	38	-0.0208 -		ISM02.4	PPB	12/11/2019 10:58	SMPL
K6169-03	MC0BW4S	4325	1.0528 -		ISM02.4	PPB	12/11/2019 11:00	SMPL
K6169-04	MC0BW5	58	-0.0158 -		ISM02.4	PPB	12/11/2019 11:03	SMPL
K6169-05	MC0BX4	71	-0.0125 -		ISM02.4	PPB	12/11/2019 11:05	SMPL
K6169-06	MC0BX5	35	-0.0215 -		ISM02.4	PPB	12/11/2019 11:07	SMPL
K6169-07	MC0BX7	35	-0.0215 -		ISM02.4	PPB	12/11/2019 11:09	SMPL
K6169-08	MC0BX8	36	-0.0213 -		ISM02.4	PPB	12/11/2019 11:12	SMPL
K6169-09	MC0BX9	45	-0.019 -		ISM02.4	PPB	12/11/2019 11:14	SMPL
K6169-10	MC0BY0	27	-0.0235 -		ISM02.4	PPB	12/11/2019 11:16	SMPL
K6169-11	MC0BY1	2	-0.0298 -		ISM02.4	PPB	12/11/2019 11:18	SMPL
K6169-12	MC0BY2	65	-0.014 -		ISM02.4	PPB	12/11/2019 11:21	SMPL
K6169-13	MC0BY3	36	-0.0213 -		ISM02.4	PPB	12/11/2019 11:23	SMPL
K6169-14	MC0BY4	48	-0.0183 -		ISM02.4	PPB	12/11/2019 11:25	SMPL
K6169-15	MC0BY6	28	-0.0233 -		ISM02.4	PPB	12/11/2019 11:27	SMPL
K6169-16	MC0BY7	24	-0.0243 -		ISM02.4	PPB	12/11/2019 11:30	SMPL
K6169-17	MC0BY8	26	-0.0238 -		ISM02.4	PPB	12/11/2019 11:32	SMPL
K6169-18	MC0BZ1	21	-0.025 -		ISM02.4	PPB	12/11/2019 11:34	SMPL
K6169-19	MC0BZ4	13	-0.027 -		ISM02.4	PPB	12/11/2019 11:37	SMPL
CCV002	CCV002	21007	5.2306 -		ISM02.4	PPB	12/11/2019 11:39	SMPL
CCB002	CCB002	-10	-0.0328 -		ISM02.4	PPB	12/11/2019 11:41	SMPL
K6169-20	MC0BZ7	30	-0.0228 -		ISM02.4	PPB	12/11/2019 11:43	SMPL
K6169-21	MC0C01	34	-0.0218 -		ISM02.4	PPB	12/11/2019 11:46	SMPL
K6169-22	MC0C03	12	-0.0273 -		ISM02.4	PPB	12/11/2019 11:48	SMPL
PB125383E	PBW383	39	-0.0205 -		ISM02.4	PPB	12/11/2019 11:50	SMPL
K6201-01	MC0BR7	43	-0.0195 -		ISM02.4	PPB	12/11/2019 11:52	SMPL
K6201-02	MC0BR8	40	-0.0203 -		ISM02.4	PPB	12/11/2019 11:55	SMPL
K6201-03	MC0BS3	48	-0.0183 -		ISM02.4	PPB	12/11/2019 11:57	SMPL
K6201-04	MC0BS4	48	-0.0183 -		ISM02.4	PPB	12/11/2019 11:59	SMPL
K6201-05	MC0BS5	41	-0.02 -		ISM02.4	PPB	12/11/2019 12:02	SMPL
K6201-06	MC0BS7	36	-0.0213 -		ISM02.4	PPB	12/11/2019 12:04	SMPL
K6201-07	MC0BT3	28	-0.0233 -		ISM02.4	PPB	12/11/2019 12:06	SMPL
K6201-08	MC0BT4	55	-0.0165 -		ISM02.4	PPB	12/11/2019 12:08	SMPL
K6201-09	MC0BT5	18	-0.0258 -		ISM02.4	PPB	12/11/2019 12:11	SMPL
K6201-10	MC0BT8	42	-0.0198 -		ISM02.4	PPB	12/11/2019 12:13	SMPL

LB106680
INSTRUMENT ID CV1

K6201-11	MC0BT9	42	-0.0198 -	ISM02.4	PPB	12/11/2019 12:15 SMPL
K6201-12	MC0BW0	33	-0.022 -	ISM02.4	PPB	12/11/2019 12:17 SMPL
K6201-13	MC0BW3	38	-0.0208 -	ISM02.4	PPB	12/11/2019 12:20 SMPL
K6201-14	MC0BW3D	0	-0.0303 -	ISM02.4	PPB	12/11/2019 12:22 SMPL
K6201-15	MC0BW3S	4901	1.1971 -	ISM02.4	PPB	12/11/2019 12:24 SMPL
K6201-16	MC0BW6	44	-0.0193 -	ISM02.4	PPB	12/11/2019 12:26 SMPL
CCV003	CCV003	20515	5.1074 -	ISM02.4	PPB	12/11/2019 12:29 SMPL
CCB003	CCB003	10	-0.0278 -	ISM02.4	PPB	12/11/2019 12:31 SMPL
K6201-17	MC0BX2	40	-0.0203 -	ISM02.4	PPB	12/11/2019 12:33 SMPL
K6201-18	MC0BX3	52	-0.0173 -	ISM02.4	PPB	12/11/2019 12:36 SMPL
K6201-19	MC0BZ0	58	-0.0158 -	ISM02.4	PPB	12/11/2019 12:38 SMPL
K6201-20	MC0BZ3	59	-0.0155 -	ISM02.4	PPB	12/11/2019 12:40 SMPL
K6201-21	MC0BZ6	193	0.0181 -	ISM02.4	PPB	12/11/2019 12:42 SMPL
K6201-22	MC0C00	93	-0.007 -	ISM02.4	PPB	12/11/2019 12:45 SMPL
CCV004	CCV004	20830	5.1862 -	ISM02.4	PPB	12/11/2019 12:47 SMPL
CCB004	CCB004	-4	-0.0313 -	ISM02.4	PPB	12/11/2019 12:49 SMPL