



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106682

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/10/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	12/10/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/10/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	12/10/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	12/10/2019	17:06
6	K5664-01	1	Solid	20.03	20	22.0	8.31	12/10/2019	17:08
7	K5664-02	1	Solid	20.04	20	21.9	8.23	12/10/2019	17:10
8	K6242-09	1	Solid	20.02	20	21.3	7.28	12/10/2019	17:11
9	K6242-10	1	Solid	20.03	20	21.4	7.58	12/10/2019	17:12
10	K6242-11	1	Solid	20.04	20	21.1	7.55	12/10/2019	17:15
11	K6242-12	1	Solid	20.03	20	21.0	8.35	12/10/2019	17:16
12	K6247-03	1	Solid	20.04	20	21.3	8.66	12/10/2019	17:19
13	K6247-06	1	Solid	20.03	20	21.6	6.99	12/10/2019	17:20
14	K6247-06DUP	1	Solid	20.04	20	21.7	7.00	12/10/2019	17:24
15	CCV2	1	Water	NA	NA	20.1	12.02	12/10/2019	17:25

WORKLIST(Hardcopy Internal Chain)

LB106682

WorkList Name : CORROSIONITY K6242

WorkList ID : 134060

Department : Wet-Chemistry

Date : 12-10-2019 16:25:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2019	Solid	K5664-01	Corrosivity	Cool 4 deg C	PSEG01	J33	01-02-04-05-06-07-COMP	12/10/2019	9045D
12/10/2019	Solid	K5664-02	Corrosivity	Cool 4 deg C	PSEG01	J33	13-14-15-16-17-18-19-COMP	12/10/2019	9045D
12/09/2019	Solid	K6242-09	Corrosivity	Cool 4 deg C	LIRO01	B41	SB-01-COMP	12/09/2019	9045D
12/09/2019	Solid	K6242-10	Corrosivity	Cool 4 deg C	LIRO01	B41	SB-02-COMP	12/09/2019	9045D
12/09/2019	Solid	K6242-11	Corrosivity	Cool 4 deg C	LIRO01	B41	SB-03-COMP	12/09/2019	9045D
12/09/2019	Solid	K6242-12	Corrosivity	Cool 4 deg C	LIRO01	B41	SB-04-COMP	12/09/2019	9045D
12/10/2019	Solid	K6247-03	Corrosivity	Cool 4 deg C	PSEG01	J32	RBR200059	12/10/2019	9045D
12/10/2019	Solid	K6247-06	Corrosivity	Cool 4 deg C	PSEG01	J32	RT3645	12/10/2019	9045D

Date/Time 12-11-14 9:45
 Received by: Jo
 Relinquished by: cd

Date/Time 12-11-19 11:00
 Received by: en
 Relinquished by: So