



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106684

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/11/2019	10:00
2	CAL2	1	Water	NA	NA	20.1	7.00	12/11/2019	10:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/11/2019	10:04
4	ICV	1	Water	NA	NA	20.2	7.01	12/11/2019	10:05
5	CCV1	1	Water	NA	NA	20.1	2.01	12/11/2019	10:06
6	K5664-04	1	Water	NA	NA	21.6	7.29	12/11/2019	10:07
7	K5664-04DUP	1	Water	NA	NA	21.7	7.30	12/11/2019	10:08
8	K5674-01	1	Water	NA	NA	16.3	6.85	12/11/2019	10:10
9	CCV2	1	Water	NA	NA	20.2	12.01	12/11/2019	10:15

WORKLIST(Hardcopy Internal Chain)

2810687

WorkList Name : PH K5674

WorkList ID : 134083

Department : Wet-Chemistry

Date : 12-11-2019 09:45:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2019	Water	K5664-04	pH	Cool 4 deg C	PSEG01	J33	02-03-08-09-10-11-12-COMP	12/10/2019	9040C
12/10/2019	Water	K5674-01	pH	Cool 4 deg C	PSEG01	J22	OILY-WATER-DRUM	12/10/2019	9040C

Date/Time 12-11-19 9:45
 Received by: JA
 Relinquished by: JA

Date/Time 12-11-19 11:00
 Received by: ew
 Relinquished by: JA