

IC-2 300.0

12-11-19 AK

Clear table


ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-02_	12/2/19 10:56		CT/AP
STD2	0.2720	0.4060	0.3910	1.2920	0.3260	0.1450	2.2400	2019-12-02_	12/2/19 11:25		CT/AP
STD3	0.3960	0.6620	0.6250	2.1250	0.5300	0.8360	3.4100	2019-12-02_	12/2/19 11:54		CT/AP
STD4	0.8280	1.2340	1.1520	3.9940	1.0030	2.1520	6.0950	2019-12-02_	12/2/19 12:23		CT/AP
STD5	1.9260	2.8420	2.6260	9.0000	2.2820	5.1270	13.8610	2019-12-02_	12/2/19 12:53		CT/AP
STD6	3.8710	5.8910	5.5700	18.7720	4.7480	9.8150	28.9730	2019-12-02_	12/2/19 13:22		CT/AP
STD7	5.1060	7.7170	7.7300	25.7100	6.4440	12.5700	38.6390	2019-12-02_	12/2/19 13:51		CT/AP
ICV	1.9790	2.9590	2.9160	9.9090	2.4610	5.4570	14.5220	2019-12-02_	12/2/19 15:18		CT/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-02_	12/2/19 15:48		CT/AP
CCV	2.0870	2.6230	2.7450	9.1410	2.3570	4.6260	13.9840	2019-12-10_	12/10/19 12:33		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-10_	12/10/19 13:14		CT/AP
LB106696BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-10_	12/10/19 13:43		CT/AP
LB106696BSW	2.0650	2.6300	2.7530	9.1780	2.3680	4.8980	13.9790	2019-12-10_	12/10/19 14:12		CT/AP
K6217-03	0.0000	110.4250	0.0000	0.0000	0.2990	0.0170	2.1090	2019-12-10_	12/10/19 14:56		CT/AP
K6218-02	0.0000	538.2370	0.0000	0.0000	1.2640	0.0000	11.9370	2019-12-10_	12/10/19 15:25		CT/AP
K6218-02DUP	0.0000	537.9470	0.0000	0.0000	1.2600	0.0000	11.8990	2019-12-10_	12/10/19 15:54		CT/AP
K6218-02MS	2.5300	459.3280	0.3410	9.7590	3.6730	10.2770	25.5090	2019-12-10_	12/10/19 16:23		CT/AP
K6218-02MSD	2.4230	462.4310	0.3420	9.4560	3.6150	10.2550	25.0550	2019-12-10_	12/10/19 16:52		CT/AP
CCV	2.1240	2.6750	2.7100	9.2400	2.4030	5.9450	13.8830	2019-12-10_	12/10/19 17:22		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-12-10_	12/10/19 17:54		CT/AP

CALIBRATION OF COMPONENT F-

IC-2 cal

Method: AnionsIC2-120219.mtw
Equation: $Q = 0.0752034 \cdot A + 1.81152$
RSD: 4.391 %
Correlation coefficient: 0.999185

12-02-19
AK



K3 = 0 K2 = 0 K1 = 0.0752034 K0 = 1.81152

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	2.296	44.86	0.2	20	5.516		
3	4.329	80.87	0.4	20	5.516		
4	8.943	195	0.8	20	5.516		
5	24.22	488.7	2	20	5.516		
6	53.88	1010	4	20	5.516		
7	71.73	1334	5	20	5.516		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.120863 \cdot A + 1.2447$
 RSD: 6.200 %
 Correlation coefficient: 0.998374




K3 = 0 K2 = 0 K1 = 0.120863 K0 = 1.2447
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.158	55.35	0.3	20	8.443		
3	5.833	99.99	0.6	20	8.443		
4	11.23	193.8	1.2	20	8.443		
5	26.29	453.2	3	20	8.443		
6	56.28	947	6	20	8.443		
7	75.7	1267	7.5	20	8.443		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0576239 \cdot A + 2.62969$
 RSD: 6.907 %
 Correlation coefficient: 0.997982



K3 = 0 K2 = 0 K1 = 0.0576239 K0 = 2.62969

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.575	104.3	0.3	20	10.18		
3	10.16	190.2	0.6	20	10.18		
4	19.71	377.2	1.2	20	10.18		
5	46.07	909.8	3	20	10.18		
6	94.95	1963	6	20	10.18		
7	122.7	2637	7.5	20	10.18		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.30011 \cdot A + 4.35517$
 RSD: 6.082 %
 Correlation coefficient: 0.998435




K3 = 0 K2 = 0 K1 = 0.30011 K0 = 4.35517
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0	0	0		0
2	3.041		71.85	1	20		13.07
3	5.607		131.9	2	20		13.07
4	11.13		262.5	4	20		13.07
5	26.41		608.2	10	20		13.07
6	56.7		1271	20	20		13.07
7	76.24		1699	25	20		13.07

±□¿\$

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0512792 \cdot A + 1.26972$
 RSD: 6.621 %
 Correlation coefficient: 0.998146




K3 = 0 K2 = 0 K1 = 0.0512792 K0 = 1.26972
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.627	106.8	0.25	20	15.04		
3	6.583	190.8	0.5	20	15.04		
4	12.97	378.3	1	20	15.04		
5	30.31	884	2.5	20	15.04		
6	64.26	1848	5	20	15.04		
7	86.84	2489	6.25	20	15.04		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.0895479 \cdot A - 10.5374$
 RSD: 4.024 %
 Correlation coefficient: 0.999316



K3 = 0 K2 = 0 K1 = 0.0895479 K0 = -10.5374

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.632	182.8	0.5	20	20.37		
3	6.723	334	1	20	20.37		
4	12.66	614.6	2	20	20.37		
5	27.3	1271	5	20	20.37		
6	51.79	2303	10	20	20.37		
7	66.92	2925	12.5	20	20.37		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-120219.mtw
 Equation: $Q = 0.180193 \cdot A + 9.63188$
 RSD: 6.659 %
 Correlation coefficient: 0.998124



K3 = 0 K2 = 0 K1 = 0.180193 K0 = 9.63188
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	4.314	178.7	1.5	20	23.43		
3	7.703	321	3	20	23.43		
4	15.05	625.1	6	20	23.43		
5	35.66	1486	15	20	23.43		
6	76.51	3155	30	20	23.43		
7	103.2	4235	37.5	20	23.43		

Report date: 12/4/2019 10:47:14 AM
Printed by: wet


Ident: STD1
Analysis from: 12/2/2019 10:56:36 AM
File: _2019-12-02_

Last save: 12/2/2019 11:23:07 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23817

Last save: 12/2/2019 9:08:

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:28 AM
Printed by: wet


Ident: STD2
Analysis from: 12/2/2019 11:25:43 AM
File: _2019-12-02_

Last save: 12/2/2019 11:52:13 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23818

Last save: 12/2/2019 11:23

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.233	2.29	8.95	44.857	6.02
2	8.35	0.205	3.16	12.31	55.349	7.43
3	10.09	0.246	5.58	21.74	104.339	14.01
4	13.02	0.301	3.04	11.86	71.853	9.65
5	14.99	0.354	3.62	14.13	106.768	14.34
6	19.95	0.649	3.63	14.16	182.847	24.55
7	22.95	0.573	4.31	16.81	178.743	24.00
7	26.50	0.366	25.63	99.95	744.757	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:42 AM
Printed by: wet


Ident: STD3
Analysis from: 12/2/2019 11:54:51 AM
File: _2019-12-02_

Last save: 12/2/2019 12:21:21 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23819

Last save: 12/2/2019 11:52

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.223	4.33	9.22	80.870	6.00
2	8.34	0.204	5.83	12.43	99.989	7.41
3	10.06	0.245	10.16	21.65	190.168	14.10
4	12.99	0.298	5.61	11.94	131.936	9.78
5	14.94	0.350	6.58	14.02	190.757	14.14
6	19.93	0.639	6.72	14.32	333.952	24.76
7	22.93	0.572	7.70	16.41	320.970	23.80
7	26.50	0.362	46.94	99.99	1348.642	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:47:54 AM
Printed by: wet


Ident: STD4
Analysis from: 12/2/2019 12:23:59 PM
File: _2019-12-02_

Last save: 12/2/2019 12:50:29 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23820

Last save: 12/2/2019 12:21

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.225	8.94	9.75	194.998	7.37
2	8.33	0.204	11.23	12.25	193.786	7.32
3	10.06	0.251	19.71	21.49	377.229	14.25
4	12.96	0.296	11.14	12.14	262.482	9.92
5	14.90	0.350	12.97	14.14	378.281	14.29
6	19.92	0.625	12.65	13.80	614.631	23.22
7	22.93	0.570	15.05	16.41	625.148	23.62
7	26.50	0.360	91.68	100.00	2646.555	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:48:40 AM
Printed by: wet


Ident: STD5
Analysis from: 12/2/2019 12:53:08 PM
File: _2019-12-02_

Last save: 12/2/2019 1:19:38 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23821

Last save: 12/2/2019 12:50

SAMPLE:
: CT/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.219	24.22	11.20	488.743	8.01
2	8.33	0.204	26.29	12.16	453.221	7.43
3	10.04	0.255	46.07	21.30	909.837	14.91
4	12.94	0.292	26.41	12.21	608.236	9.97
5	14.83	0.351	30.31	14.02	884.022	14.49
6	19.90	0.599	27.30	12.62	1270.514	20.82
7	22.93	0.565	35.66	16.49	1486.471	24.36
7	26.50	0.355	216.25	100.00	6101.043	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:48:56 AM
 Printed by: wet


Ident: STD6
 Analysis from: 12/2/2019 1:22:16 PM
 File: _2019-12-02_

Last save: 12/2/2019 1:48:47 PM

Method: AnionsIC2-120219.mtw
 Run operator: wet
 Analysis number: 23822

Last save: 12/2/2019 1:19:

SAMPLE:
 : CT/AP
 Vial number: 7
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.207	53.88	11.86	1009.951	8.08
2	8.32	0.200	56.28	12.39	947.013	7.58
3	10.01	0.264	94.94	20.89	1962.950	15.71
4	12.90	0.286	56.70	12.48	1271.405	10.17
5	14.76	0.347	64.26	14.14	1848.352	14.79
6	19.86	0.573	51.78	11.40	2302.756	18.43
7	22.91	0.555	76.51	16.84	3155.054	25.25
7	26.50	0.347	454.36	100.00	12497.482	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 12/4/2019 10:49:05 AM
Printed by: wet


Ident: STD7
Analysis from: 12/2/2019 1:51:25 PM
File: _2019-12-02_

Last save: 12/2/2019 2:17:55 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23823

Last save: 12/2/2019 1:48:

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.207	71.73	11.89	1333.776	8.04
2	8.33	0.199	75.70	12.55	1266.716	7.64
3	10.00	0.277	122.69	20.34	2637.313	15.90
4	12.90	0.285	76.24	12.64	1698.834	10.24
5	14.74	0.346	86.83	14.39	2488.596	15.00
6	19.86	0.564	66.92	11.09	2925.006	17.64
7	22.92	0.552	103.17	17.10	4235.156	25.54
7	26.50	0.347	603.27	100.00	16585.396	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:50:23 AM
Printed by: wet


Ident: ICV
Analysis from: 12/2/2019 3:18:52 PM
File: _2019-12-02_

Last save: 12/4/2019 10:50:14 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23826

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.217	24.75	10.96	502.172	7.82
2	8.33	0.204	27.81	12.31	479.323	7.46
3	10.04	0.261	48.00	21.25	966.604	15.05
4	12.94	0.293	27.84	12.33	645.858	10.05
5	14.83	0.353	31.90	14.12	935.064	14.56
6	19.91	0.603	28.58	12.65	1336.367	20.80
7	22.94	0.572	37.01	16.38	1558.420	24.26
7	26.50	0.358	225.90	100.00	6423.808	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/4/2019 10:50:33 AM
Printed by: wet


Ident: ICB
Analysis from: 12/2/2019 3:48:02 PM
File: _2019-12-02_

Last save: 12/4/2019 10:50:15 AM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23827

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:15:58 PM
Printed by: wet


Ident: CCV
Analysis from: 12/10/2019 12:33:34 PM
File: _2019-12-10_

Last save: 12/10/2019 1:00:05 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23913

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.41	0.170	37.03	15.91	530.860	8.85
2	8.32	0.205	26.35	11.32	423.746	7.06
3	10.01	0.256	47.18	20.27	907.144	15.12
4	12.90	0.293	26.73	11.49	594.680	9.91
5	14.78	0.360	30.93	13.29	894.375	14.91
6	19.82	0.574	25.78	11.07	1150.814	19.18
7	22.71	0.525	38.74	16.64	1498.638	24.98
7	26.50	0.340	232.75	100.00	6000.258	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:16:07 PM
Printed by: wet


Ident: CCB
Analysis from: 12/10/2019 1:14:24 PM
File: _2019-12-10_

Last save: 12/10/2019 1:40:55 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23914

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:16:16 PM
Printed by: wet


Ident: LB106696BLW
Analysis from: 12/10/2019 1:43:33 PM
File: _2019-12-10_

Last save: 12/11/2019 4:15:50 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23915

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:16:24 PM
Printed by: wet


Ident: LB106696BSW
Analysis from: 12/10/2019 2:12:41 PM
File: _2019-12-10_

Last save: 12/11/2019 4:15:50 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23916

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.43	0.188	31.75	13.88	525.131	8.66
2	8.31	0.204	26.54	11.60	424.942	7.01
3	10.01	0.258	46.94	20.52	909.717	15.00
4	12.88	0.292	26.95	11.78	597.143	9.84
5	14.76	0.358	31.19	13.64	898.883	14.82
6	19.84	0.578	26.98	11.80	1211.531	19.97
7	22.73	0.530	38.40	16.78	1498.154	24.70
7	26.50	0.344	228.75	100.00	6065.500	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:16:36 PM
Printed by: wet


Ident: K6217-03
Analysis from: 12/10/2019 2:56:22 PM
File: _2019-12-10_

Last save: 12/10/2019 3:22:52 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23917

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.39	0.201	1155.87	99.17	18262.535	97.89
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	14.89	0.361	2.94	0.25	91.674	0.49
6	19.96	0.744	2.18	0.19	121.388	0.65
7	22.75	0.548	4.59	0.39	180.604	0.97
7	26.50	0.265	1165.59	100.00	18656.201	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:16:52 PM
Printed by: wet


Ident: K6218-02
Analysis from: 12/10/2019 3:25:30 PM
File: _2019-12-10_

Last save: 12/10/2019 3:52:01 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23918

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.83	0.792	1656.44	97.20	89055.317	98.08
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	14.83	0.353	15.72	0.92	468.177	0.52
6	0.00	0.000	0.00	0.00	0.000	0.00
7	22.76	0.539	31.94	1.87	1271.494	1.40
7	26.50	0.241	1704.10	100.00	90794.988	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:17:55 PM
Printed by: wet


Ident: K6218-02DUP
Analysis from: 12/10/2019 3:54:38 PM
File: _2019-12-10_

Last save: 12/10/2019 4:21:09 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23919

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 10
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.83	0.791	1657.13	97.22	89007.334	98.09
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	14.84	0.354	15.61	0.92	466.731	0.51
6	0.00	0.000	0.00	0.00	0.000	0.00
7	22.77	0.539	31.82	1.87	1267.276	1.40
7	26.50	0.241	1704.56	100.00	90741.341	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:46:03 PM
Printed by: wet


Ident: K6218-02MS
Analysis from: 12/10/2019 4:23:47 PM
File: _2019-12-10_

Last save: 12/11/2019 4:46:01 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23920

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.357	22.24	1.21	648.668	0.77
2	8.79	0.735	1605.73	87.67	75997.675	90.52
3	10.81	0.277	4.31	0.24	72.580	0.09
4	12.93	0.287	29.08	1.59	635.840	0.76
5	14.77	0.356	48.19	2.63	1407.972	1.68
6	19.89	0.629	49.02	2.68	2412.965	2.87
7	22.76	0.526	73.01	3.99	2777.856	3.31
7	26.50	0.452	1831.59	100.00	83953.556	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:46:20 PM
Printed by: wet


Ident: K6218-02MSD
Analysis from: 12/10/2019 4:52:55 PM
File: _2019-12-10_

Last save: 12/11/2019 4:46:17 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23921

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.46	0.355	21.14	1.16	620.318	0.74
2	8.80	0.738	1606.90	87.90	76511.230	90.72
3	10.83	0.282	4.25	0.23	73.030	0.09
4	12.94	0.286	28.19	1.54	615.634	0.73
5	14.78	0.355	47.27	2.59	1385.232	1.64
6	19.89	0.631	48.72	2.67	2407.977	2.86
7	22.77	0.527	71.58	3.92	2727.469	3.23
7	26.50	0.454	1828.07	100.00	84340.890	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:46:29 PM
Printed by: wet


Ident: CCV
Analysis from: 12/10/2019 5:22:07 PM
File: _2019-12-10_

Last save: 12/10/2019 5:48:38 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23922

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.45	0.225	25.50	11.36	540.901	8.57
2	8.32	0.206	26.75	11.92	432.428	6.85
3	10.03	0.262	45.56	20.29	894.868	14.17
4	12.90	0.296	26.80	11.94	601.229	9.52
5	14.79	0.362	31.14	13.87	912.355	14.45
6	19.88	0.600	30.83	13.73	1445.433	22.89
7	22.78	0.538	37.92	16.89	1487.455	23.56
7	26.50	0.356	224.48	100.00	6314.669	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/11/2019 4:47:04 PM
Printed by: wet


Ident: CCB
Analysis from: 12/10/2019 5:54:38 PM
File: _2019-12-10_

Last save: 12/10/2019 6:21:09 PM

Method: AnionsIC2-120219.mtw
Run operator: wet
Analysis number: 23923

Last save: 12/2/2019 2:17:

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD