



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106701

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/11/2019	09:35
2	K5664-05		1	8	100.00	>212.0	>212.0	12/11/2019	09:55
3	K5664-05DUP		1	8	100.00	>212.0	>212.0	12/11/2019	10:20
4	K5674-02		1	9	100.00	>212.0	>212.0	12/11/2019	10:45
5	K6179-01		1	7	60.00	140.0	140.0	12/11/2019	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP121119 WorkList ID : 134080 Department : Wet-Chemistry Date : 12-11-2019 08:59:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/23/2019	Water	K5664-05	Flash Point	Cool 4 deg C	PSEG01	J33	02-03-08-09-10-11-12-COMP	12/10/2019	1010A
12/17/2019	Water	K5674-02	Flash Point	Cool 4 deg C	PSEG01	J22	OILY-WATER-DRUM	12/10/2019	1010A
12/19/2019	Water	K6179-01	Flash Point	Cool 4 deg C	WARR01	H31	UC-11-26-19	12/05/2019	1010A

Date/Time 12/11/19 09:05
 Received by: KS
 Relinquished by: SP

Date/Time 12/11/19 11:30
 Received by: SP
 Relinquished by: KS