


LB106705

INSTRUMENT 1D CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	194	0.000	194				
0.2	0.200	0.214	0.014	1495	0.0 %	1495				
2.5	2.500	2.473	-0.027	14705	0.0 %	14705				
5.0	5.000	5.032	0.032	29664	0.0 %	29664				
7.5	7.500	7.501	0.001	44102	0.0 %	44102				
10.0	10.000	9.990	-0.010	58654	0.0 %	58654				

LB106705 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 12 Dec 2019 07:40:03				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	194	-	0.0000	7471B	PPB	12 Dec 2019 07:52:08	S Std
0.2 - 1	50.2	1495	-	0.2000	7471B	PPB	12 Dec 2019 07:54:25	S Std
2.5 - 1	52.5	14705	-	2.5000	7471B	PPB	12 Dec 2019 07:56:42	S Std
5.0 - 1	55.0	29664	-	5.0000	7471B	PPB	12 Dec 2019 07:58:59	S Std
7.5 - 1	57.5	44102	-	7.5000	7471B	PPB	12 Dec 2019 08:03:29	S Std
10.0 - 1	510.0	58654	-	10.0000	7471B	PPB	12 Dec 2019 08:09:12	S Std
ICV25 - 1	ICV25	23643	4.0017	-	7471B	PPB	12 Dec 2019 08:11:54	U SMPL
ICB25 - 1	ICB25	57	-0.0323	-	7471B	PPB	12 Dec 2019 08:14:10	U SMPL
CCV05 - 1	CCV05	29712	5.0397	-	7471B	PPB	12 Dec 2019 08:16:25	U SMPL
CCB05 - 1	CCB05	39	-0.0354	-	7471B	PPB	12 Dec 2019 08:18:40	U SMPL
CRA - 1	CRA	1550	0.2230	-	7471B	PPB	12 Dec 2019 08:20:56	U SMPL
HighStd - 1	HighStd	58518	9.9666	-	7471B	PPB	12 Dec 2019 08:23:11	U SMPL
ChkStd - 1	ChkStd	41144	6.9950	-	7471B	PPB	12 Dec 2019 08:27:10	U SMPL
PB125402BL - 1	PBS	42	-0.0349	-	7471B	PPB	12 Dec 2019 08:29:26	U SMPL
PB125402BS - 1	LCSS	23757	4.0212	-	7471B	PPB	12 Dec 2019 08:33:27	U SMPL
K5664-01 - 1	01-02-04-05-06-07-COMP	11783	1.9732	-	7471B	PPB	12 Dec 2019 08:35:44	U SMPL
K5664-02 - 1	13-14-15-16-17-18-19-COMP	212	-0.0058	-	7471B	PPB	12 Dec 2019 08:38:01	U SMPL
K5792-01 - 1	1	1035	0.1350	-	7471B	PPB	12 Dec 2019 08:40:18	U SMPL
K6204-06MS - 1	K072-AA-WC-1	50672	8.6246	-	7471B	PPB	12 Dec 2019 08:42:35	U SMPL
K6204-06MSD - 1	K072-AA-WC-1DUP	50846	8.6544	-	7471B	PPB	12 Dec 2019 08:44:51	U SMPL
CCV06 - 1	CCV06	29731	5.0430	-	7471B	PPB	12 Dec 2019 08:47:08	U SMPL
CCB06 - 1	CCB06	10	-0.0404	-	7471B	PPB	12 Dec 2019 08:49:25	U SMPL
K6204-06 - 1	K072-AA-WC-1MS	36877	6.2652	-	7471B	PPB	12 Dec 2019 08:51:41	U SMPL
K6204-06DUP - 1	K072-AA-WC-1MSD	36647	6.2259	-	7471B	PPB	12 Dec 2019 08:53:56	U SMPL
K6204-07 - 1	K072-AA-WC-2	416	0.0291	-	7471B	PPB	12 Dec 2019 08:56:13	U SMPL
K6204-13 - 1	K072-AB-WC-1	2271	0.3464	-	7471B	PPB	12 Dec 2019 08:58:29	U SMPL
K6205-06 - 1	K165-WC-1	5917	0.9699	-	7471B	PPB	12 Dec 2019 09:00:45	U SMPL
K6205-07 - 1	K165-WC-2	737	0.0840	-	7471B	PPB	12 Dec 2019 09:03:02	U SMPL
K6245-01 - 1	OUTSIDE-BRICK-1	220	-0.0044	-	7471B	PPB	12 Dec 2019 09:05:19	U SMPL
K6245-03 - 1	OUTSIDE-BRICK-2	145	-0.0173	-	7471B	PPB	12 Dec 2019 09:07:36	U SMPL
K6245-05 - 1	DARK-GREEN-BRICK	246	0.0000	-	7471B	PPB	12 Dec 2019 09:09:54	U SMPL
K6245-07 - 1	LIGHT-GREEN-BRICK	126	-0.0205	-	7471B	PPB	12 Dec 2019 09:12:10	U SMPL
CCV07 - 1	CCV07	30075	5.1018	-	7471B	PPB	12 Dec 2019 09:14:25	U SMPL
CCB07 - 1	CCB07	206	-0.0068	-	7471B	PPB	12 Dec 2019 09:16:41	U SMPL
K6245-09 - 1	MEZZ-FLOOR	249	0.0005	-	7471B	PPB	12 Dec 2019 09:18:56	U SMPL
K6245-11 - 1	RELAY-FLOOR	1585	0.2290	-	7471B	PPB	12 Dec 2019 09:21:11	U SMPL
K6245-13 - 1	WHITE-BRICK	81	-0.0282	-	7471B	PPB	12 Dec 2019 09:23:27	U SMPL
K6245-16 - 1	BASEMENT-FLOOR	13	-0.0398	-	7471B	PPB	12 Dec 2019 09:25:43	U SMPL
K6247-01 - 1	RBR200059	2325	0.3556	-	7471B	PPB	12 Dec 2019 09:27:57	U SMPL
K6247-04 - 1	RT3465	898	0.1115	-	7471B	PPB	12 Dec 2019 09:30:13	U SMPL
K6204-06LX5 - 1		7354	1.2157	-	7471B	PPB	12 Dec 2019 09:37:28	U SMPL
K6204-06A - 1		56673	9.6510	-	7471B	PPB	12 Dec 2019 09:46:40	U SMPL
CCV08 - 1	CCV08	29606	5.0216	-	7471B	PPB	12 Dec 2019 09:48:58	U SMPL
CCB08 - 1	CCB08	193	-0.0091	-	7471B	PPB	12 Dec 2019 09:51:15	U SMPL