



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106719

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/12/2019	14:00
2	CAL2	1	Water	NA	NA	20.1	7.00	12/12/2019	14:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/12/2019	14:02
4	ICV	1	Water	NA	NA	20.1	7.00	12/12/2019	14:05
5	CCV1	1	Water	NA	NA	20.1	2.02	12/12/2019	14:06
6	K6269-06	1	Water	NA	NA	17.6	7.30	12/12/2019	14:09
7	K6269-06DUP	1	Water	NA	NA	17.7	7.30	12/12/2019	14:10
8	CCV2	1	Water	NA	NA	20.1	12.02	12/12/2019	14:11

2B 106719

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6269 WorkList ID : 134138 Department : Wet-Chemistry Date : 12-12-2019 13:46:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2019	Water	K6269-06	pH	Cool 4 deg C	PSEG01	H12	23810	12/12/2019	9040C

Date/Time 12.12.19 13.50
Received by: JD
Relinquished by: cd

Date/Time 12.12.19 14.30
Received by: JD
Relinquished by: JD