



PERCENT SOLID

Supervisor: apatel
Analyst: ketankumar
Date: 12/13/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 12/12/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:15
Out Date: 12/13/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106726

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K6110-01	HD-02-121219	1	1.14	9.96	8.33	81.5	
K6110-02	HD-02-121219	2	1.16	9.94	8.72	86.1	
K6110-03	HD-02-121219-1	3	1.14	9.96	8.33	81.5	
K6267-01	WO-3	4	1.11	9.71	8.54	86.4	
K6267-02	WO-4	5	1.15	9.94	8.3	81.3	
K6268-01	33-19-12-11-01E	6	1.12	9.88	8.79	87.6	
K6268-02	33-19-12-11-01W	7	1.12	9.66	8.5	86.4	
K6271-01	BE-213-COMP-S	8	1.15	9.85	8.47	84.1	
K6274-01	T2-COMP	9	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6274-02	T2-COMP	10	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6274-03	T2-COMP	11	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6283-01	SOIL-CUTTINGS	12	1.12	9.94	8.31	81.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$