


LB106732
INSTRUMENT ID 001

ISM02.4



Linear

A= 0.0000e+000

B= 2.6273e-004 *y-intercept*

C= -7.8229e-002 *slope*

Rho= 0.9991251

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.056	-0.056	84	0.000	84				
0.05	0.050									
0.20	0.200	0.225	0.025	1156	0.0 %	1156				
2.50	2.500	2.347	-0.153	9231	0.0 %	9231				
5.00	5.000	5.288	0.288	20426	0.0 %	20426				
7.50	7.500	7.564	0.064	29087	0.0 %	29087				
10.0	10.000	9.831	-0.169	37717	0.0 %	37717				

0/0

*.12
-6
6
0
-2*

LB106732
INSTRUMENT ID CV1

Sample ID	Extended II μ Abs.	Conc.	Std C	Method	Units	Date	Type
	0 S0.0	84 -		0 ISM02.4	PPB	12/13/2019 7:44	Std
	0.2 S0.2	1156 -		0.2 ISM02.4	PPB	12/13/2019 7:46	Std
	2.5 S2.5	9231 -		2.5 ISM02.4	PPB	12/13/2019 7:49	Std
	5 S5.0	20426 -		5 ISM02.4	PPB	12/13/2019 7:51	Std
	7.5 S7.5	29087 -		7.5 ISM02.4	PPB	12/13/2019 7:53	Std
	10 S10.0	37717 -		10 ISM02.4	PPB	12/13/2019 7:58	Std
ICV	ICV	15987	4.1221 -	ISM02.4	PPB	12/13/2019 8:03	SMPL
ICB	ICB	11	-0.0753 -	ISM02.4	PPB	12/13/2019 8:05	SMPL
CCV013	CCV013	20641	5.3449 -	ISM02.4	PPB	12/13/2019 8:07	SMPL
CCB013	CCB013	3	-0.0774 -	ISM02.4	PPB	12/13/2019 8:09	SMPL
PB125435BL	PBW435	16	-0.074 -	ISM02.4	PPB	12/13/2019 8:12	SMPL
K6207-01	MC0C02	53	-0.0643 -	ISM02.4	PPB	12/13/2019 8:14	SMPL
K6207-02	MC0C28	42	-0.0672 -	ISM02.4	PPB	12/13/2019 8:16	SMPL
K6207-03	MC0C32	92	-0.0541 -	ISM02.4	PPB	12/13/2019 8:18	SMPL
K6207-04	MC0C36	52	-0.0646 -	ISM02.4	PPB	12/13/2019 8:21	SMPL
K6207-05	MC0C38	36	-0.0688 -	ISM02.4	PPB	12/13/2019 8:23	SMPL
K6207-06	MC0C42	90	-0.0546 -	ISM02.4	PPB	12/13/2019 8:25	SMPL
K6207-07	MC0C44	69	-0.0601 -	ISM02.4	PPB	12/13/2019 8:28	SMPL
K6207-08	MC0C49	60	-0.0625 -	ISM02.4	PPB	12/13/2019 8:30	SMPL
K6207-09	MC0C51	62	-0.0619 -	ISM02.4	PPB	12/13/2019 8:32	SMPL
K6207-10	MC0C51D	59	-0.0627 -	ISM02.4	PPB	12/13/2019 8:34	SMPL
K6207-11	MC0C51S	4555	1.1185 -	ISM02.4	PPB	12/13/2019 8:37	SMPL
PB125436BL	PBW436	34	-0.0693 -	ISM02.4	PPB	12/13/2019 8:39	SMPL
K6213-01	MC0BY5	75	-0.0585 -	ISM02.4	PPB	12/13/2019 8:41	SMPL
K6213-02	MC0C08	45	-0.0664 -	ISM02.4	PPB	12/13/2019 8:43	SMPL
K6213-03	MC0C10	38	-0.0682 -	ISM02.4	PPB	12/13/2019 8:46	SMPL
K6213-04	MC0C13	38	-0.0682 -	ISM02.4	PPB	12/13/2019 8:48	SMPL
K6213-05	MC0C17	47	-0.0659 -	ISM02.4	PPB	12/13/2019 8:50	SMPL
K6213-06	MC0C19	30	-0.0703 -	ISM02.4	PPB	12/13/2019 8:52	SMPL
K6213-07	MC0C23	79	-0.0575 -	ISM02.4	PPB	12/13/2019 8:55	SMPL
CCV014	CCV014	19073	4.9329 -	ISM02.4	PPB	12/13/2019 9:00	SMPL
CCB014	CCB014	23	-0.0722 -	ISM02.4	PPB	12/13/2019 9:03	SMPL
K6213-08	MC0C25	42	-0.0672 -	ISM02.4	PPB	12/13/2019 9:05	SMPL
K6213-09	MC0C29	43	-0.0669 -	ISM02.4	PPB	12/13/2019 9:07	SMPL
K6213-10	MC0C33	32	-0.0698 -	ISM02.4	PPB	12/13/2019 9:09	SMPL
K6213-11	MC0C37	56	-0.0635 -	ISM02.4	PPB	12/13/2019 9:12	SMPL
K6213-12	MC0C39	39	-0.068 -	ISM02.4	PPB	12/13/2019 9:14	SMPL
K6213-13	MC0C43	59	-0.0627 -	ISM02.4	PPB	12/13/2019 9:16	SMPL
K6213-14	MC0C45	46	-0.0661 -	ISM02.4	PPB	12/13/2019 9:18	SMPL
K6213-15	MC0C50	48	-0.0656 -	ISM02.4	PPB	12/13/2019 9:21	SMPL
K6213-16	MC0C52	50	-0.0651 -	ISM02.4	PPB	12/13/2019 9:23	SMPL
K6213-17	MC0C52D	43	-0.0669 -	ISM02.4	PPB	12/13/2019 9:25	SMPL
K6213-18	MC0C52S	4851	1.1963 -	ISM02.4	PPB	12/13/2019 9:28	SMPL
CCV015	CCV015	19429	5.0264 -	ISM02.4	PPB	12/13/2019 9:30	SMPL
CCB015	CCB015	-32	-0.0866 -	ISM02.4	PPB	12/13/2019 9:32	SMPL