


LB106764
INSTRUMENT ID CV1

7471B



Linear

A= 0.0000e+000

B= 1.8180e-004

C= -3.6902e-002

Rho= 0.9996901

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.039	-0.039	-11	0.000	-11				
0.2	0.200	0.169	-0.031	1135	0.0 %	1135				
2.5	2.500	2.479	-0.021	13839	0.0 %	13839				
5.0	5.000	5.203	0.203	28822	0.0 %	28822				
7.5	7.500	7.452	-0.048	41193	0.0 %	41193				
10.0	10.000	9.935	-0.065	54850	0.0 %	54850				

LB106764 INSTRUMENT ID CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Dec 2019 08:27:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-11	-	0.0000	7471B	PPB	16 Dec 2019 08:35:29	S	Std
0.2 - 1	50.2	1135	-	0.2000	7471B	PPB	16 Dec 2019 08:37:46	S	Std
2.5 - 1	52.5	13839	-	2.5000	7471B	PPB	16 Dec 2019 08:40:04	S	Std
5.0 - 1	55.0	28822	-	5.0000	7471B	PPB	16 Dec 2019 08:42:20	S	Std
7.5 - 1	57.5	41193	-	7.5000	7471B	PPB	16 Dec 2019 08:44:37	S	Std
10.0 - 1	510.0	54850	-	10.0000	7471B	PPB	16 Dec 2019 08:46:54	S	Std
ICV30 - 1	ICV30	21822	3.9304	-	7471B	PPB	16 Dec 2019 08:49:56	U	SMPL
ICB30 - 1	ICB30	-60	-0.0478	-	7471B	PPB	16 Dec 2019 08:52:11	U	SMPL
CCV23 - 1	CCV23	29723	5.3669	-	7471B	PPB	16 Dec 2019 08:54:27	U	SMPL
CCB23 - 1	CCB23	14	-0.0344	-	7471B	PPB	16 Dec 2019 08:57:17	U	SMPL
CRA - 1	CRA	1272	0.1944	-	7471B	PPB	16 Dec 2019 08:59:33	U	SMPL
HighStd - 1	HighStd	56086	10.1598	-	7471B	PPB	16 Dec 2019 09:01:48	U	SMPL
ChkStd - 1	ChkStd	38912	7.0375	-	7471B	PPB	16 Dec 2019 09:04:03	U	SMPL
PB125479BL - 1	PBS	-44	-0.0449	-	7471B	PPB	16 Dec 2019 09:06:19	U	SMPL
PB125479BS - 1	LCSS	22591	4.0703	-	7471B	PPB	16 Dec 2019 09:14:18	U	SMPL
K6110-01 - 1	HD-02-121219	2216	0.3660	-	7471B	PPB	16 Dec 2019 09:16:34	U	SMPL
K6110-01DUP - 1	HD-02-121219DUP	2303	0.3818	-	7471B	PPB	16 Dec 2019 09:18:51	U	SMPL
K6110-01MS - 1	HD-02-121219MS	22161	3.9921	-	7471B	PPB	16 Dec 2019 09:24:54	U	SMPL
K6110-01MSD - 1	HD-02-121219MSD	22303	4.0179	-	7471B	PPB	16 Dec 2019 09:27:11	U	SMPL
K6112-05 - 1	OK-03-121219	2604	0.4365	-	7471B	PPB	16 Dec 2019 09:29:27	U	SMPL
CCV24 - 1	CCV24	27065	4.8836	-	7471B	PPB	16 Dec 2019 09:31:42	U	SMPL
CCB24 - 1	CCB24	-17	-0.0400	-	7471B	PPB	16 Dec 2019 09:33:58	U	SMPL
K6265-03 - 1	TP-1A-S	18003	3.2361	-	7471B	PPB	16 Dec 2019 09:36:13	U	SMPL
K6265-07 - 1	TP-1A-D	7056	1.2459	-	7471B	PPB	16 Dec 2019 09:38:29	U	SMPL
K6267-01 - 1	WO-3	2640	0.4431	-	7471B	PPB	16 Dec 2019 09:40:44	U	SMPL
K6267-02 - 1	WO-4	3321	0.5669	-	7471B	PPB	16 Dec 2019 09:44:31	U	SMPL
K6274-01 - 1	T2-COMP	7	-0.0356	-	7471B	PPB	16 Dec 2019 09:46:47	U	SMPL
K6284-01 - 1	42-SPRUCE	37332	6.7502	-	7471B	PPB	16 Dec 2019 09:49:03	U	SMPL
K6284-02 - 1	48-SPRUCE	16789	3.0154	-	7471B	PPB	16 Dec 2019 09:51:20	U	SMPL
K6110-01LX5 - 1		437	0.0425	-	7471B	PPB	16 Dec 2019 09:53:38	U	SMPL
CCV25 - 1	CCV25	28086	5.0693	-	7471B	PPB	16 Dec 2019 09:55:56	U	SMPL
CCB25 - 1	CCB25	-21	-0.0407	-	7471B	PPB	16 Dec 2019 09:58:11	U	SMPL