



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106766

Slope : 99.2

BalanceID: WC SC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/16/2019	11:05
2	CAL2	1	Water	NA	NA	20.2	7.01	12/16/2019	11:06
3	CAL3	1	Water	NA	NA	20.2	10.03	12/16/2019	11:07
4	ICV	1	Water	NA	NA	20.2	7.00	12/16/2019	11:09
5	CCV1	1	Water	NA	NA	20.3	2.02	12/16/2019	11:10
6	K6309-01	1	Solid	20.03	20	21.5	9.33	12/16/2019	11:11
7	K6309-01DUP	1	Solid	20.04	20	21.6	9.34	12/16/2019	11:12
8	K6309-02	1	Solid	20.02	20	21.7	9.51	12/16/2019	11:15
9	K6309-03	1	Solid	20.03	20	22.2	6.07	12/16/2019	11:16
10	K6309-04	1	Solid	20.04	20	21.2	9.70	12/16/2019	11:19
11	K6309-05	1	Solid	20.02	20	21.1	9.05	12/16/2019	11:20
12	K6309-06	1	Solid	20.03	20	21.1	6.70	12/16/2019	11:22
13	K6309-07	1	Solid	20.02	20	22.5	8.83	12/16/2019	11:24
14	CCV2	1	Water	NA	NA	20.3	12.02	12/16/2019	11:25

WORKLIST(Hardcopy Internal Chain)

106766

WorkList Name : PH K6309 WorkList ID : 134231 Department : Wet-Chemistry Date : 12-16-2019 10:52:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2019	Solid	K6309-01	pH	Cool 4 deg C	PSEG01	H13	30784	12/13/2019	9045D
12/13/2019	Solid	K6309-02	pH	Cool 4 deg C	PSEG01	H13	30785	12/13/2019	9045D
12/13/2019	Solid	K6309-03	pH	Cool 4 deg C	PSEG01	H13	30786	12/13/2019	9045D
12/13/2019	Solid	K6309-04	pH	Cool 4 deg C	PSEG01	H13	30787	12/13/2019	9045D
12/13/2019	Solid	K6309-05	pH	Cool 4 deg C	PSEG01	H13	30788	12/13/2019	9045D
12/13/2019	Solid	K6309-06	pH	Cool 4 deg C	PSEG01	H13	30789	12/13/2019	9045D
12/13/2019	Solid	K6309-07	pH	Cool 4 deg C	PSEG01	H13	30790	12/13/2019	9045D

Date/Time 12-16-19 11:00
 Received by: Jr
 Relinquished by: ed

Date/Time 12-16-19 12:20
 Received by: ed
 Relinquished by: Jr