



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106768

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/16/2019	09:15
2	K6258-01		1	8	100.00	>212.0	>212.0	12/16/2019	09:40
3	K6258-01DUP		1	8	100.00	>212.0	>212.0	12/16/2019	10:10
4	K6258-02		1	9	100.00	>212.0	>212.0	12/16/2019	10:35
5	K6269-06		1	9	100.00	>212.0	>212.0	12/16/2019	10:55
6	K6283-02		1	8	100.00	>212.0	>212.0	12/16/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121319 WorkList ID : 134183 Department : Wet-Chemistry Date : 12-13-2019 15:13:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2019	Water	K6258-01	Flash Point	Cool 4 deg C	DIST01	B41	MH-NEAR-THE-CURB-(3A)	12/11/2019	1010A
12/18/2019	Water	K6258-02	Flash Point	Cool 4 deg C	DIST01	B41	MH-NEAR-THE-FRONT-DOO	12/11/2019	1010A
12/23/2019	Water	K6269-06	Flash Point	Cool 4 deg C	PSEG01	H12	23810	12/12/2019	1010A
12/26/2019	Water	K6283-02	Flash Point	Cool 4 deg C	SEAC01	B61	GROUNDWATER	12/12/2019	1010A

Date/Time 12/16/19 08:45
 Received by: KS
 Relinquished by: JP

Date/Time 12/16/19 11:25
 Received by: JP
 Relinquished by: KS