


LB106777

INSTRUMENT ID CV1

245.1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	33	0.000	33				
0.05	0.050	0.037	-0.013	201	0.0 %	201				
0.2	0.200	0.153	-0.047	648	0.0 %	648				
2.5	2.500	2.442	-0.058	9453	0.0 %	9453				
5.0	5.000	5.245	0.245	20231	0.0 %	20231				
7.5	7.500	7.482	-0.018	28834	0.0 %	28834				
10.0	10.000	9.899	-0.101	38131	0.0 %	38131				

LB106777 INSTRUMENT ID CV1

Method: 245.1

Operator: Admin

Date of Analysis: 16 Dec 2019 14:34:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC-0	33	-	0.0000	245.1	PPB	16 Dec 2019 14:35:59	S	Std
0.05 - 1	SC-05	201	-	0.0500	245.1	PPB	16 Dec 2019 14:57:07	S	Std
0.2 - 1	SC-2	648	-	0.2000	245.1	PPB	16 Dec 2019 14:59:24	S	Std
2.5 - 1	SC-6	9453	-	2.5000	245.1	PPB	16 Dec 2019 15:01:41	S	Std
5.0 - 1	SC-0	20231	-	5.0000	245.1	PPB	16 Dec 2019 15:03:58	S	Std
7.5 - 1	SC-5	28834	-	7.5000	245.1	PPB	16 Dec 2019 15:06:14	S	Std
10.0 - 1	SC-0	38131	-	10.0000	245.1	PPB	16 Dec 2019 15:08:30	S	Std
ICV31 - 1	ICV31	16240	4.2070	-245.1		PPB	16 Dec 2019 15:22:17	U	SMPL
ICB31 - 1	ICB31	14	-0.0121	-245.1		PPB	16 Dec 2019 15:24:32	U	SMPL
CCV26 - 1	CCV26	20395	5.2874	-245.1		PPB	16 Dec 2019 15:26:48	U	SMPL
CCB26 - 1	CCB26	41	-0.0051	-245.1		PPB	16 Dec 2019 15:29:03	U	SMPL
CRA - 1	CRA	173	0.0292	-245.1		PPB	16 Dec 2019 15:38:19	U	SMPL
HighStd - 1	HighStd	39087	10.1477	-245.1		PPB	16 Dec 2019 15:43:57	U	SMPL
ChkStd - 1	ChkStd	25549	6.6276	-245.1		PPB	16 Dec 2019 15:46:13	U	SMPL
PB125499BL - 1	PBW	-16	-0.0199	-245.1		PPB	16 Dec 2019 15:48:28	U	SMPL
PB125499BS - 1	LCSW	14475	3.7481	-245.1		PPB	16 Dec 2019 15:50:45	U	SMPL
K6258-01 - 1	MH-NEAR-THE-CURB-(3A)	113	0.0136	-245.1		PPB	16 Dec 2019 15:54:34	U	SMPL
K6258-01DUP - 1	MH-NEAR-THE-CURB-(3A)DUP	111	0.0131	-245.1		PPB	16 Dec 2019 15:56:50	U	SMPL
K6258-01MS - 1	MH-NEAR-THE-CURB-(3A)MS	20133	5.2193	-245.1		PPB	16 Dec 2019 15:59:07	U	SMPL
K6258-01MSD - 1	MH-NEAR-THE-CURB-(3A)MSD	20217	5.2411	-245.1		PPB	16 Dec 2019 16:01:25	U	SMPL
K6258-02 - 1	MH-NEAR-THE-FRONT-DOOR-(4A)	58	-0.0007	-245.1		PPB	16 Dec 2019 16:03:40	U	SMPL
CCV27 - 1	CCV27	18162	4.7068	-245.1		PPB	16 Dec 2019 16:17:44	U	SMPL
CCB27 - 1	CCB27	-12	-0.0189	-245.1		PPB	16 Dec 2019 16:19:59	U	SMPL
K6312-08 - 1	EFF-WW	-5	-0.0170	-245.1		PPB	16 Dec 2019 16:22:13	U	SMPL
K6315-03 - 1	NCQ-632	4	-0.0147	-245.1		PPB	16 Dec 2019 16:24:27	U	SMPL
K6316-03 - 1	OB-722	51	-0.0025	-245.1		PPB	16 Dec 2019 16:26:41	U	SMPL
K6321-02 - 1	EFFLUENT-METALS	288	0.0591	-245.1		PPB	16 Dec 2019 16:28:57	U	SMPL
K6258-01LX5 - 1		8	-0.0137	-245.1		PPB	16 Dec 2019 16:31:13	U	SMPL
K6258-01A - 1		14273	3.6956	-245.1		PPB	16 Dec 2019 16:33:30	U	SMPL
CCV28 - 1	CCV28	19804	5.1337	-245.1		PPB	16 Dec 2019 16:35:46	U	SMPL
CCB28 - 1	CCB28	-30	-0.0235	-245.1		PPB	16 Dec 2019 16:38:04	U	SMPL