



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106785

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/16/2019	17:35
2	CAL2	1	Water	NA	NA	20.1	7.00	12/16/2019	17:36
3	CAL3	1	Water	NA	NA	20.1	10.03	12/16/2019	17:37
4	ICV	1	Water	NA	NA	20.1	7.01	12/16/2019	17:38
5	CCV1	1	Water	NA	NA	20.2	2.01	12/16/2019	17:39
6	K6331-02	1	Water	NA	NA	18.5	6.99	12/16/2019	17:40
7	K6331-02DUP	1	Water	NA	NA	18.6	7.00	12/16/2019	17:41
8	CCV2	1	Water	NA	NA	20.1	12.01	12/16/2019	17:42

WORKLIST(Hardcopy Internal Chain)

18106785

WorkList Name : CORROSIONITY K6331 WorkList ID : 134257 Department : Wet-Chemistry Date : 12-16-2019 17:26:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2019	Water	K6331-02	☒ Corrosivity	Cool 4 deg C	PSEG01	H21	N48353	12/16/2019	9040C

Date/Time 12-16-19 17:36
 Received by: Ja
 Relinquished by: ed

Date/Time 12-16-19 18:00
 Received by: ed
 Relinquished by: Ja