



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106786

Slope : 98.8

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2568         |
| buffer solution pH 7 yellow    | W2447         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 12/16/2019 | 17:35     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 12/16/2019 | 17:36     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 12/16/2019 | 17:37     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 12/16/2019 | 17:38     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 12/16/2019 | 17:39     |
| 6   | K6331-01    | 1  | Water  | NA          | NA          | 18.5             | 6.99        | 12/16/2019 | 17:40     |
| 7   | K6331-01DUP | 1  | Water  | NA          | NA          | 18.6             | 7.00        | 12/16/2019 | 17:41     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.01       | 12/16/2019 | 17:42     |

2B 106786

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6331      WorkList ID : 134258      Department : Wet-Chemistry      Date : 12-16-2019 17:26:50

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 12/16/2019 | Water  | K6331-01 | pH   | Cool 4 deg C | PSEG01   | H21              | N48353          | 12/16/2019   | 9040C  |

Date/Time 12-16-19 17:30  
Received by: [Signature]  
Relinquished by: [Signature]

Date/Time 12-16-19 18:00  
Received by: [Signature]  
Relinquished by: [Signature]