


LB106789
INSTRUMENT ID 01

ISM02.4



Linear

A= 0.0000e+000

B= 2.4238e-004

C= 5.6699e-002

Rho= 0.9995584

Accept=Accepted

12/17/19

slope
y-intercept

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.070	0.070	54	0.000	54				
0.05	0.050									
0.20	0.200	0.236	0.036	740	0.0 %	740				
2.50	2.500	2.424	-0.076	9767	0.0 %	9767				
5.00	5.000	5.015	0.015	20458	0.0 %	20458				
7.50	7.500	7.306	-0.194	29908	0.0 %	29908				
10.0	10.000	10.149	0.149	41639	0.0 %	41639				

1. D

13
-3
0
-3
1

12/17/19

LB106789
INSTRUMENT ID CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	54	-	0	ISM02.4	PPB	12/17/2019 9:45	Std
0.2 S0.2	740	-	0.2	ISM02.4	PPB	12/17/2019 9:48	Std
2.5 S2.5	9767	-	2.5	ISM02.4	PPB	12/17/2019 9:50	Std
5 S5.0	20458	-	5	ISM02.4	PPB	12/17/2019 9:52	Std
7.5 S7.5	29908	-	7.5	ISM02.4	PPB	12/17/2019 9:54	Std
10 S10.0	41639	-	10	ISM02.4	PPB	12/17/2019 9:57	Std
ICV	ICV	15777	3.8807 -	ISM02.4	PPB	12/17/2019 10:00	SMPL
ICB	ICB	-10	0.0543 -	ISM02.4	PPB	12/17/2019 10:02	SMPL
CCV033	CCV033	20424	5.007 -	ISM02.4	PPB	12/17/2019 10:06	SMPL
CCB033	CCB033	-21	0.0516 -	ISM02.4	PPB	12/17/2019 10:08	SMPL
PB125497BL	PBW497	13	0.0598 -	ISM02.4	PPB	12/17/2019 10:10	SMPL
K6207-01	MCOC02	52	0.0693 -	ISM02.4	PPB	12/17/2019 10:13	SMPL
K6207-02	MCOC28	42	0.0669 -	ISM02.4	PPB	12/17/2019 10:15	SMPL
K6207-03	MCOC32	60	0.0712 -	ISM02.4	PPB	12/17/2019 10:17	SMPL
K6207-04	MCOC36	43	0.0671 -	ISM02.4	PPB	12/17/2019 10:19	SMPL
K6207-05	MCOC38	34	0.0649 -	ISM02.4	PPB	12/17/2019 10:22	SMPL
K6207-06	MCOC42	86	0.0775 -	ISM02.4	PPB	12/17/2019 10:24	SMPL
K6207-07	MCOC44	59	0.071 -	ISM02.4	PPB	12/17/2019 10:26	SMPL
K6207-08	MCOC49	56	0.0703 -	ISM02.4	PPB	12/17/2019 10:29	SMPL
K6207-09	MCOC51	64	0.0722 -	ISM02.4	PPB	12/17/2019 10:31	SMPL
K6207-10	MCOC82	31	0.0642 -	ISM02.4	PPB	12/17/2019 10:33	SMPL
K6207-13	MCOC86	4	0.0577 -	ISM02.4	PPB	12/17/2019 10:35	SMPL
K6207-14	MCOC88	36	0.0654 -	ISM02.4	PPB	12/17/2019 10:38	SMPL
K6207-15	MCOC90	38	0.0659 -	ISM02.4	PPB	12/17/2019 10:40	SMPL
K6207-16	MCOC90S	4534	1.1556 -	ISM02.4	PPB	12/17/2019 10:42	SMPL
K6207-17	MCOC90D	28	0.0635 -	ISM02.4	PPB	12/17/2019 10:44	SMPL
PB125498BL	PBW498	40	0.0664 -	ISM02.4	PPB	12/17/2019 10:47	SMPL
K6213-01	MCOCBY5	42	0.0669 -	ISM02.4	PPB	12/17/2019 10:49	SMPL
K6213-02	MCOC08	39	0.0662 -	ISM02.4	PPB	12/17/2019 10:51	SMPL
K6213-03	MCOC10	37	0.0657 -	ISM02.4	PPB	12/17/2019 10:53	SMPL
CCV034	CCV034	20925	5.1285 -	ISM02.4	PPB	12/17/2019 10:56	SMPL
CCB034	CCB034	-13	0.0535 -	ISM02.4	PPB	12/17/2019 10:58	SMPL
K6213-04	MCOC13	5	0.0579 -	ISM02.4	PPB	12/17/2019 11:00	SMPL
K6213-05	MCOC17	-7	0.055 -	ISM02.4	PPB	12/17/2019 11:02	SMPL
K6213-06	MCOC19	37	0.0657 -	ISM02.4	PPB	12/17/2019 11:05	SMPL
K6213-07	MCOC23	47	0.0681 -	ISM02.4	PPB	12/17/2019 11:07	SMPL
K6213-08	MCOC25	1	0.0569 -	ISM02.4	PPB	12/17/2019 11:09	SMPL
K6213-09	MCOC29	33	0.0647 -	ISM02.4	PPB	12/17/2019 11:11	SMPL
K6213-10	MCOC33	31	0.0642 -	ISM02.4	PPB	12/17/2019 11:14	SMPL
K6213-11	MCOC37	51	0.0691 -	ISM02.4	PPB	12/17/2019 11:16	SMPL
K6213-12	MCOC39	11	0.0594 -	ISM02.4	PPB	12/17/2019 11:18	SMPL
K6213-13	MCOC43	42	0.0669 -	ISM02.4	PPB	12/17/2019 11:20	SMPL
K6213-14	MCOC45	47	0.0681 -	ISM02.4	PPB	12/17/2019 11:23	SMPL
K6213-15	MCOC50	53	0.0695 -	ISM02.4	PPB	12/17/2019 11:25	SMPL
K6213-16	MCOC52	29	0.0637 -	ISM02.4	PPB	12/17/2019 11:27	SMPL
K6213-17	MCOC83	31	0.0642 -	ISM02.4	PPB	12/17/2019 11:30	SMPL

12/17/19

LB106789
INSTRUMENT ID CV1

K6213-18	MC0C87	31	0.0642 -	ISM02.4	PPB	12/17/2019 11:32 SMPL
K6213-19	MC0C89	54	0.0698 -	ISM02.4	PPB	12/17/2019 11:34 SMPL
K6213-20	MC0C91	29	0.0637 -	ISM02.4	PPB	12/17/2019 11:36 SMPL
K6213-21	MC0C91D	35	0.0652 -	ISM02.4	PPB	12/17/2019 11:39 SMPL
K6213-22	MC0C91S	4921	1.2494 - ✓	ISM02.4	PPB	12/17/2019 11:41 SMPL
PB125514BL	PBW514	24	0.0625 - ✓	ISM02.4	PPB	12/17/2019 11:43 SMPL
CCV035	CCV035	21131	5.1784 - ✓	ISM02.4	PPB	12/17/2019 11:45 SMPL
CCB035	CCB035	-9	0.0545 - ✓	ISM02.4	PPB	12/17/2019 11:48 SMPL
K6304-01	MC0C63	550	0.19 -	ISM02.4	PPB	12/17/2019 11:50 SMPL
K6304-02	MC0C73	18	0.0611 -	ISM02.4	PPB	12/17/2019 11:52 SMPL
K6304-03	MC0C75	33	0.0647 -	ISM02.4	PPB	12/17/2019 11:55 SMPL
K6304-04	MC0C75D	41	0.0666 -	ISM02.4	PPB	12/17/2019 11:57 SMPL
K6304-05	MC0C75S	4602	1.1721 - ✓	ISM02.4	PPB	12/17/2019 11:59 SMPL
K6304-06	MC0C78	56	0.0703 -	ISM02.4	PPB	12/17/2019 12:01 SMPL
K6304-07	MC0C56	1071	0.3163 -	ISM02.4	PPB	12/17/2019 12:04 SMPL
K6304-08	MC0C61	28	0.0635 -	ISM02.4	PPB	12/17/2019 12:06 SMPL
K6304-09	MC0C65	216	0.1091 -	ISM02.4	PPB	12/17/2019 12:08 SMPL
PB125515BL	PBW515	8	0.0586 - ✓	ISM02.4	PPB	12/17/2019 12:10 SMPL
K6305-01	MC0C64	23	0.0623 -	ISM02.4	PPB	12/17/2019 12:13 SMPL
K6305-02	MC0C74	32	0.0645 -	ISM02.4	PPB	12/17/2019 12:15 SMPL
K6305-03	MC0C76	20	0.0615 -	ISM02.4	PPB	12/17/2019 12:17 SMPL
K6305-04	MC0C76D	27	0.0632 -	ISM02.4	PPB	12/17/2019 12:20 SMPL
K6305-05	MC0C76S	4771	1.2131 - ✓	ISM02.4	PPB	12/17/2019 12:22 SMPL
K6305-06	MC0C79	15	0.0603 -	ISM02.4	PPB	12/17/2019 12:24 SMPL
K6305-07	MC0C57	20	0.0615 -	ISM02.4	PPB	12/17/2019 12:26 SMPL
K6305-08	MC0C62	24	0.0625 -	ISM02.4	PPB	12/17/2019 12:29 SMPL
K6305-09	MC0C66	36	0.0654 -	ISM02.4	PPB	12/17/2019 12:31 SMPL
CCV036	CCV036	21543	5.2783 - ✓	ISM02.4	PPB	12/17/2019 12:33 SMPL
CCB036	CCB036	-6	0.0552 - ✓	ISM02.4	PPB	12/17/2019 12:35 SMPL