



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/18/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 12/17/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:00
Out Date: 12/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106796

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6277-01	C0C27	1	1.12	9.91	4.84	42.3	
K6277-02	C0C34	2	1.14	9.78	6.57	62.8	
K6277-03	C0C35	3	1.16	9.95	6.37	59.3	
K6277-04	C0C40	4	1.12	9.56	5.79	55.3	
K6277-05	C0C46	5	1.12	9.98	6.58	61.6	
K6277-06	C0C60	6	1.15	9.52	6.46	63.4	
K6277-07	C0C67	7	1.16	9.87	6.07	56.4	
K6277-08	C0C68	8	1.15	9.71	5.12	46.4	
K6277-09	C0C69	9	1.14	9.89	4.44	37.7	
K6277-10	C0C70	10	1.15	9.87	6.77	64.4	
K6277-11	C0C71	11	1.15	9.5	7.71	78.6	
K6277-12	C0C72	12	1.12	9.82	8.38	83.4	
K6277-13	C0C48	13	1.14	9.8	6.8	65.4	
K6277-14	C0C53	14	1.13	9.89	6.81	64.8	
K6277-15	C0C54	15	1.12	9.8	6.03	56.6	
K6277-16	C0C55	16	1.12	9.9	7.03	67.3	
K6277-17	C0C58	17	1.13	9.5	4.77	43.5	
K6277-18	C0C59	18	1.14	9.53	6.4	62.7	
K6277-19	VHBLK01	19	1.00	2.00	2.00	100.0	VHBLK
K6277-20	C0C93	20	1.13	9.7	6.45	62.1	
K6277-21	C0C93MS	21	1.13	9.7	6.45	62.1	
K6277-22	C0C93MSD	22	1.13	9.7	6.45	62.1	
K6277-23	C0C80	23	1.14	9.55	5.89	56.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$