



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106799

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	12/17/2019	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	12/17/2019	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/17/2019	15:02
4	ICV	1	Water	NA	NA	20.1	7.00	12/17/2019	15:05
5	CCV1	1	Water	NA	NA	20.1	2.02	12/17/2019	15:06
6	K6329-01	1	Solid	20.03	20	20.7	6.20	12/17/2019	15:07
7	K6329-01DUP	1	Solid	20.04	20	20.8	6.21	12/17/2019	15:08
8	CCV2	1	Water	NA	NA	20.1	12.02	12/17/2019	15:10

13106799

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6329 WorkList ID : 134306 Department : Wet-Chemistry Date : 12-17-2019 14:43:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2019	Solid	K6329-01	pH	Cool 4 deg C	PSEG01	H21	121619-CA-COMP	12/16/2019	9045D

Date/Time 12-17-19 14:45
Received by: JA
Relinquished by: JA

Date/Time 12-17-19 15:45
Received by: cm
Relinquished by: Jo