



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106807

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/17/2019	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	12/17/2019	14:11
3	CAL3	1	Water	NA	NA	20.1	10.02	12/17/2019	14:12
4	ICV	1	Water	NA	NA	20.1	7.00	12/17/2019	14:15
5	CCV1	1	Water	NA	NA	20.1	2.01	12/17/2019	14:16
6	K6353-02	1	Water	NA	NA	14.9	7.23	12/17/2019	14:17
7	K6353-02DUP	1	Water	NA	NA	14.8	7.24	12/17/2019	14:19
8	K6353-03	1	Water	NA	NA	16.6	7.12	12/17/2019	14:20
9	K6353-04	1	Water	NA	NA	17.5	7.43	12/17/2019	14:21
10	K6353-05	1	Water	NA	NA	18.2	7.48	12/17/2019	14:22
11	K6355-01	1	Water	NA	NA	16.6	6.39	12/17/2019	14:24
12	K6355-02	1	Water	NA	NA	18.1	8.32	12/17/2019	14:25
13	CCV2	1	Water	NA	NA	20.2	12.01	12/17/2019	14:26

287 106807

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6353 WorkList ID : 134300 Department : Wet-Chemistry Date : 12-17-2019 13:58:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2019	Water	K6353-02	✓ pH	Cool 4 deg C	TETR06	D61	TT-012-IDWGW-20191216	12/16/2019	9040C
12/16/2019	Water	K6353-03	✓ pH	Cool 4 deg C	TETR06	D61	TT-013-IDWGW-20191216	12/16/2019	9040C
12/16/2019	Water	K6353-04	✓ pH	Cool 4 deg C	TETR06	D61	TT-014-IDWGW-20191216	12/16/2019	9040C
12/16/2019	Water	K6353-05	✓ pH	Cool 4 deg C	TETR06	D61	TT-015-IDWGW-20191216	12/16/2019	9040C
12/17/2019	Water	K6355-01	✓ pH	Cool 4 deg C	PSEG01	H21	30604	12/17/2019	9040C
12/17/2019	Water	K6355-02	✓ pH	Cool 4 deg C	PSEG01	H21	13954	12/17/2019	9040C

Date/Time 12-17-19 14:00
 Received by: JP
 Relinquished by: JP

Date/Time 12-17-19 16:30
 Received by: JP
 Relinquished by: JP