



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106808

Slope : 99.1

BalanceID: WC SC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/17/2019	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	12/17/2019	14:11
3	CAL3	1	Water	NA	NA	20.1	10.02	12/17/2019	14:12
4	ICV	1	Water	NA	NA	20.1	7.00	12/17/2019	14:15
5	CCV1	1	Water	NA	NA	20.1	2.01	12/17/2019	16:40
6	K6359-03	1	Solid	20.03	20	21.5	10.40	12/17/2019	16:41
7	K6359-04	1	Solid	20.05	20	20.9	10.30	12/17/2019	16:43
8	K6359-05	1	Solid	20.02	20	20.8	10.98	12/17/2019	16:45
9	K6360-02	1	Solid	20.02	20	20.3	7.40	12/17/2019	16:46
10	K6360-02DUP	1	Solid	20.01	20	20.4	7.41	12/17/2019	16:47
11	CCV2	1	Water	NA	NA	20.1	12.02	12/17/2019	16:49

106808

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K6359 WorkList ID : 134307 Department : Wet-Chemistry Date : 12-17-2019 14:43:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2019	Solid	K6359-03	Corrosivity	Cool 4 deg C	PSEG01	H22	CH-FLOORS	12/17/2019	9045D
12/17/2019	Solid	K6359-04	Corrosivity	Cool 4 deg C	PSEG01	H22	TRANS-CON	12/17/2019	9045D
12/17/2019	Solid	K6359-05	Corrosivity	Cool 4 deg C	PSEG01	H22	FR-CON	12/17/2019	9045D
12/17/2019	Solid	K6360-02	Corrosivity	Cool 4 deg C	PSEG01	H22	1FT-TP	12/17/2019	9045D

Date/Time 12-17-19 14:45
 Received by: Jo
 Relinquished by: Jo

Date/Time 12-17-19 17:30
 Received by: Jo
 Relinquished by: Jo