



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 12/18/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/17/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 08:38
Out Date: 12/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106810

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6284-01	42-SPRUCE	1	1.15	9.95	7.78	75.3	
K6284-02	48-SPRUCE	2	1.12	9.69	7.77	77.6	
K6332-01	01-A-01-B-01-C	3	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6335-01	SB-2BD	4	1.15	9.73	9.31	95.1	
K6335-02	SB-4BD	5	1.15	9.65	9.18	94.5	
K6335-03	SB-6BD	6	1.15	9.97	9.44	94.0	
K6335-04	SB-7BD	7	1.12	9.55	9.1	94.7	
K6335-05	SB-1AD	8	1.18	9.97	9.55	95.2	
K6350-01	RO-1-3	9	1.14	9.94	9.69	97.2	
K6351-01	FL-WASTE-OIL	10	1.00	2.00	2.00	100.0	OIL SAMPLE 100% SOLIDS
K6357-01	PF-01	11	1.12	9.92	8.05	78.8	
K6357-02	PF-02	12	1.13	9.93	8.38	82.4	
K6357-03	PF-03	13	1.15	9.92	8.31	81.6	
K6357-04	PF-04	14	1.13	9.94	8.65	85.4	
K6357-05	PF-05	15	1.15	9.96	8.34	81.6	
K6357-06	PF-A	16	1.1	9.64	8.00	80.8	
K6357-07	PF-B	17	1.18	9.57	7.83	79.3	
K6357-08	PF-C	18	1.17	9.95	8.07	78.6	
K6357-09	PF-D	19	1.16	9.87	8.19	80.7	
K6357-10	PF-E	20	1.15	9.65	7.9	79.4	
K6357-11	PF-F	21	1.14	9.94	8.28	81.1	
K6357-12	PF-G	22	1.17	9.98	8.78	86.4	
K6357-13	PF-H	23	1.12	9.59	8.01	81.3	
K6357-14	PF-I	24	1.14	9.93	8.61	85.0	
K6357-15	PF-J	25	1.14	9.93	8.51	83.8	
K6357-16	PF-K	26	1.17	9.96	8.3	81.1	
K6357-17	PF-L	27	1.14	9.92	8.44	83.1	
K6357-18	PF-M	28	1.14	9.78	8.11	80.7	



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 12/18/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/17/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 08:38
Out Date: 12/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106810

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K6358-01	PCB-121619-01	29	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-02	PCB-121619-02	30	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-03	PCB-121619-03	31	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-04	PCB-121619-04	32	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-05	PCB-121619-05	33	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-06	PCB-121619-06	34	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-07	PCB-121619-07	35	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-08	PCB-121619-08	36	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6358-09	PCB-121619-09	37	1.00	2.00	2.00	100.0	CAULKING MATERIAL 100% SOLIDS
K6359-01	CH-WALLS	38	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6359-03	CH-FLOORS	39	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6359-04	TRANS-CON	40	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6359-05	FR-CON	41	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K6360-01	1FT-TP	42	1.17	9.8	8.39	83.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$