


LB106812

INSTRUMENT ID CV4

7470A



Linear

A= 0.0000e+000

B= 2.3713e-004

C= 7.8450e-002

Rho= 0.9985526

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.084	0.084	22	0.000	22				
0.2	0.200	0.250	0.050	722	0.0 %	722				
2.5	2.500	2.441	-0.059	9964	0.0 %	9964				
5.0	5.000	5.068	0.068	21041	0.0 %	21041				
7.5	7.500	7.106	-0.394	29635	0.0 %	29635				
10.0	10.000	10.252	0.252	42902	0.0 %	42902				

LB106812 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Dec 2019 07:33:56					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	22	-	0.0000	7470A	PPB	18 Dec 2019 07:34:46	S	Std
0.2 - 1	50.2	722	-	0.2000	7470A	PPB	18 Dec 2019 07:36:54	S	Std
2.5 - 1	52.6	9964	-	2.5000	7470A	PPB	18 Dec 2019 07:39:02	S	Std
5.0 - 1	56.0	21041	-	5.0000	7470A	PPB	18 Dec 2019 07:41:10	S	Std
7.5 - 1	57.5	29635	-	7.5000	7470A	PPB	18 Dec 2019 07:43:18	S	Std
10.0 - 1	59.0	42902	-	10.0000	7470A	PPB	18 Dec 2019 07:45:24	S	Std
ICV34 - 1	ICV34	15636	3.7862	-	7470A	PPB	18 Dec 2019 07:48:20	U	SMPL
ICB34 - 1	ICB34	-9	0.0763	-	7470A	PPB	18 Dec 2019 07:50:26	U	SMPL
CCV33 - 1	CCV33	20667	4.9792	-	7470A	PPB	18 Dec 2019 07:54:05	U	SMPL
CCB33 - 1	CCB33	-27	0.0720	-	7470A	PPB	18 Dec 2019 07:56:10	U	SMPL
CRA - 1	CRA	735	0.2527	-	7470A	PPB	18 Dec 2019 07:58:16	U	SMPL
HighStd - 1	HighStd	42069	10.0543	-	7470A	PPB	18 Dec 2019 08:01:53	U	SMPL
ChkStd - 1	ChkStd	29774	7.1387	-	7470A	PPB	18 Dec 2019 08:03:59	U	SMPL
PB125552BL - 1	PBW	18	0.0827	-	7470A	PPB	18 Dec 2019 08:06:05	U	SMPL
PB125552BS - 1	LCSW	15696	3.8004	-	7470A	PPB	18 Dec 2019 08:09:46	U	SMPL
K6024-10 - 1	B-1-(0.5-1.0)	6135	1.5332	-	7470A	PPB	18 Dec 2019 08:11:53	U	SMPL
K6024-11 - 1	B-2-(0.5-1.0)	45440	10.8536	-	7470A	PPB	18 Dec 2019 08:15:40	U	SMPL
K6024-12 - 1	B-3-(0.5-1.0)	26975	6.4750	-	7470A	PPB	18 Dec 2019 08:17:47	U	SMPL
K6024-13 - 1	B-4-(0.5-1.0)	129268	30.7317	-	7470A	PPB	18 Dec 2019 08:19:56	U	SMPL
K6024-14 - 1	B-5-(0.5-1.0)	34267	8.2042	-	7470A	PPB	18 Dec 2019 08:22:02	U	SMPL
CCV34 - 1	CCV34	20585	4.9598	-	7470A	PPB	18 Dec 2019 08:26:15	U	SMPL
CCB34 - 1	CCB34	20	0.0832	-	7470A	PPB	18 Dec 2019 08:28:21	U	SMPL
K6024-15 - 1	B-6-(0.5-1.0)	6940	1.7241	-	7470A	PPB	18 Dec 2019 08:30:27	U	SMPL
K6024-15DUP - 1	B-6-(0.5-1.0)DUP	6827	1.6973	-	7470A	PPB	18 Dec 2019 08:32:33	U	SMPL
K6024-15MS - 1	B-6-(0.5-1.0)MS	23013	5.5355	-	7470A	PPB	18 Dec 2019 08:34:39	U	SMPL
K6024-15MSD - 1	B-6-(0.5-1.0)MSD	22791	5.4829	-	7470A	PPB	18 Dec 2019 08:36:46	U	SMPL
PB125553BL - 1	PBW	-3	0.0777	-	7470A	PPB	18 Dec 2019 08:38:53	U	SMPL
PB125553BS - 1	LCSW	15412	3.7331	-	7470A	PPB	18 Dec 2019 08:41:00	U	SMPL
K6112-09 - 1	OK-03-121219	50	0.0903	-	7470A	PPB	18 Dec 2019 08:43:08	U	SMPL
K6112-09DUP - 1	OK-03-121219DUP	53	0.0910	-	7470A	PPB	18 Dec 2019 08:45:16	U	SMPL
K6112-09MS - 1	OK-03-121219MS	15274	3.7004	-	7470A	PPB	18 Dec 2019 08:47:24	U	SMPL
K6112-09MSD - 1	OK-03-121219MSD	15441	3.7400	-	7470A	PPB	18 Dec 2019 08:49:30	U	SMPL
CCV35 - 1	CCV35	20294	4.8908	-	7470A	PPB	18 Dec 2019 08:51:36	U	SMPL
CCB35 - 1	CCB35	37	0.0872	-	7470A	PPB	18 Dec 2019 08:53:43	U	SMPL
K6113-10 - 1	OR-02-121319	80	0.0974	-	7470A	PPB	18 Dec 2019 08:55:49	U	SMPL
K6117-06 - 1	TR-05-121319	50	0.0903	-	7470A	PPB	18 Dec 2019 08:57:55	U	SMPL
K6118-09 - 1	SU-03-121319	70	0.0950	-	7470A	PPB	18 Dec 2019 09:00:01	U	SMPL
K6294-03 - 1	TP-1	62	0.0932	-	7470A	PPB	18 Dec 2019 09:02:08	U	SMPL
K6294-06 - 1	TP-2	55	0.0915	-	7470A	PPB	18 Dec 2019 09:04:15	U	SMPL
K6310-04 - 1	SP-01	21	0.0834	-	7470A	PPB	18 Dec 2019 09:06:22	U	SMPL
PB125554BL - 1	PBW	225	0.1318	-	7470A	PPB	18 Dec 2019 09:08:29	U	SMPL
PB125554BS - 1	LCSW	15523	3.7594	-	7470A	PPB	18 Dec 2019 09:10:37	U	SMPL
K6317-09 - 1	178K-WC1	12	0.0813	-	7470A	PPB	18 Dec 2019 09:12:45	U	SMPL
K6317-09DUP - 1	178K-WC1DUP	53	0.0910	-	7470A	PPB	18 Dec 2019 09:14:51	U	SMPL
CCV36 - 1	CCV36	20215	4.8720	-	7470A	PPB	18 Dec 2019 09:19:43	U	SMPL
CCB36 - 1	CCB36	193	0.1242	-	7470A	PPB	18 Dec 2019 09:21:49	U	SMPL
K6317-09MS - 1	178K-WC1MS	15693	3.7997	-	7470A	PPB	18 Dec 2019 09:23:56	U	SMPL
K6317-09MSD - 1	178K-WC1MSD	15920	3.8536	-	7470A	PPB	18 Dec 2019 09:26:02	U	SMPL
K6317-10 - 1	178K-WC2	13	0.0815	-	7470A	PPB	18 Dec 2019 09:28:09	U	SMPL
K6318-09 - 1	X089-WC-1	38	0.0875	-	7470A	PPB	18 Dec 2019 09:30:16	U	SMPL
K6318-10 - 1	X089-WC-2	46	0.0894	-	7470A	PPB	18 Dec 2019 09:32:23	U	SMPL
K6325-03 - 1	RT-3860	40	0.0879	-	7470A	PPB	18 Dec 2019 09:34:30	U	SMPL
K6325-06 - 1	RT-3425	45	0.0891	-	7470A	PPB	18 Dec 2019 09:36:38	U	SMPL
K6329-03 - 1	121619-CA-COMP	48	0.0898	-	7470A	PPB	18 Dec 2019 09:38:46	U	SMPL
K6330-01 - 1	11978	63	0.0934	-	7470A	PPB	18 Dec 2019 09:40:54	U	SMPL
K6330-02 - 1	12013	57	0.0920	-	7470A	PPB	18 Dec 2019 09:43:01	U	SMPL
CCV37 - 1	CCV37	20868	5.0269	-	7470A	PPB	18 Dec 2019 09:45:09	U	SMPL
CCB37 - 1	CCB37	202	0.1264	-	7470A	PPB	18 Dec 2019 09:47:16	U	SMPL
K6330-03 - 1	339	50	0.0903	-	7470A	PPB	18 Dec 2019 09:49:23	U	SMPL
K6330-04 - 1	VNJ-224	48	0.0898	-	7470A	PPB	18 Dec 2019 09:51:30	U	SMPL
K6330-05 - 1	11977	60	0.0927	-	7470A	PPB	18 Dec 2019 09:53:36	U	SMPL
K6330-06 - 1	11444	53	0.0910	-	7470A	PPB	18 Dec 2019 09:55:43	U	SMPL
K6333-05 - 1	TP-1A-COMP	68	0.0946	-	7470A	PPB	18 Dec 2019 09:57:50	U	SMPL
PB125555BL - 1	PBW	18	0.0827	-	7470A	PPB	18 Dec 2019 09:59:57	U	SMPL
PB125555BS - 1	LCSW	15414	3.7336	-	7470A	PPB	18 Dec 2019 10:02:05	U	SMPL
K6269-06 - 1	23810	107	0.1038	-	7470A	PPB	18 Dec 2019 10:04:12	U	SMPL
K6331-02 - 1	N48353	53	0.0910	-	7470A	PPB	18 Dec 2019 10:06:20	U	SMPL
K6331-02DUP - 1	N48353DUP	70	0.0950	-	7470A	PPB	18 Dec 2019 10:08:27	U	SMPL
CCV38 - 1	CCV38	21235	5.1139	-	7470A	PPB	18 Dec 2019 10:10:34	U	SMPL
CCB38 - 1	CCB38	88	0.0993	-	7470A	PPB	18 Dec 2019 10:12:41	U	SMPL
K6331-02MS - 1	N48353MS	15502	3.7544	-	7470A	PPB	18 Dec 2019 10:14:48	U	SMPL
K6331-02MSD - 1	N48353MSD	15841	3.8348	-	7470A	PPB	18 Dec 2019 10:16:54	U	SMPL
PB125561BL - 1	PBW	-1	0.0782	-	7470A	PPB	18 Dec 2019 10:19:01	U	SMPL
PB125561BS - 1	LCSW	15631	3.7850	-	7470A	PPB	18 Dec 2019 10:21:08	U	SMPL
K6283-02 - 1	GROUNDWATER	232	0.1335	-	7470A	PPB	18 Dec 2019 10:23:15	U	SMPL
K6300-01 - 1	SSI-MW-4	127	0.1086	-	7470A	PPB	18 Dec 2019 10:25:22	U	SMPL
K6300-01DUP - 1	SSI-MW-4DUP	137	0.1109	-	7470A	PPB	18 Dec 2019 10:27:29	U	SMPL
K6300-02 - 1	K6300-01MS	16180	3.9152	-	7470A	PPB	18 Dec 2019 10:29:37	U	SMPL
K6300-03 - 1	K6300-01MSD	13695	3.3259	-	7470A	PPB	18 Dec 2019 10:31:45	U	SMPL
K6300-04 - 1	SSI-MW-1	164	0.1173	-	7470A	PPB	18 Dec 2019 10:33:52	U	SMPL
CCV39 - 1	CCV39	21482	5.1725	-	7470A	PPB	18 Dec 2019 10:35:59	U	SMPL

LB106812 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Dec 2019 07:33:56					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB39 - 1	CCB39	-8	0.0766	-7470A	PPB		18 Dec 2019 10:38:06	U	SMPL
K6300-05 - 1	SSI-BFD	256	0.1392	-7470A	PPB		18 Dec 2019 10:40:13	U	SMPL
K6300-06 - 1	SSI-MW-2	281	0.1451	-7470A	PPB		18 Dec 2019 10:42:20	U	SMPL
K6300-07 - 1	SSI-MW-3	315	0.1531	-7470A	PPB		18 Dec 2019 10:44:27	U	SMPL
K6331-01 - 1	N48353	221	0.1309	-7470A	PPB		18 Dec 2019 10:46:34	U	SMPL
K6353-02 - 1	TT-012-IDWGW-20191216	141	0.1119	-7470A	PPB		18 Dec 2019 10:48:41	U	SMPL
K6353-03 - 1	TT-013-IDWGW-20191216	85	0.0986	-7470A	PPB		18 Dec 2019 10:50:48	U	SMPL
K6353-04 - 1	TT-014-IDWGW-20191216	319	0.1541	-7470A	PPB		18 Dec 2019 10:52:55	U	SMPL
K6353-05 - 1	TT-015-IDWGW-20191216	448	0.1847	-7470A	PPB		18 Dec 2019 10:55:04	U	SMPL
K6024-15LX5 - 1		1535	0.4424	-7470A	PPB		18 Dec 2019 10:57:12	U	SMPL
K6112-09LX5 - 1		15	0.0820	-7470A	PPB		18 Dec 2019 10:59:19	U	SMPL
CCV40 - 1	CCV40	21345	5.1400	-7470A	PPB		18 Dec 2019 11:01:27	U	SMPL
CCB40 - 1	CCB40	-7	0.0768	-7470A	PPB		18 Dec 2019 11:03:34	U	SMPL
K6331-02LX5 - 1		3	0.0792	-7470A	PPB		18 Dec 2019 11:05:40	U	SMPL
K6317-09LX5 - 1		12	0.0813	-7470A	PPB		18 Dec 2019 11:07:47	U	SMPL
K6300-01LX5 - 1		33	0.0863	-7470A	PPB		18 Dec 2019 11:09:54	U	SMPL
PB125552TB - 1	PB125226TB	19	0.0830	-7470A	PPB		18 Dec 2019 11:12:01	U	SMPL
PB125553TB - 1	PB125469TB	-4	0.0775	-7470A	PPB		18 Dec 2019 11:14:08	U	SMPL
PB125554TB - 1	PB125506TB	18	0.0827	-7470A	PPB		18 Dec 2019 11:16:12	U	SMPL
PB125555TB - 1	PB125551TB	30	0.0856	-7470A	PPB		18 Dec 2019 11:18:20	U	SMPL
K6024-11DLX2 - 1		25246	6.0650	-7470A	PPB		18 Dec 2019 11:20:28	U	SMPL
K6024-13DLX5 - 1		28256	6.7788	-7470A	PPB		18 Dec 2019 11:22:36	U	SMPL
CCV41 - 1	CCV41	21103	5.0826	-7470A	PPB		18 Dec 2019 11:24:44	U	SMPL
CCB41 - 1	CCB41	123	0.1076	-7470A	PPB		18 Dec 2019 11:26:51	U	SMPL