



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106821

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/18/2019	11:05
2	K6355-01		1	9	100.00	>212.0	>212.0	12/18/2019	11:32
3	K6355-01DUP		1	9	100.00	>212.0	>212.0	12/18/2019	11:55
4	K6355-02		1	8	100.00	>212.0	>212.0	12/18/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 106821

WorkList Name : fp121819 WorkList ID : 134335 Department : Wet-Chemistry Date : 12-18-2019 10:28:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/24/2019	Water	K6355-01 E	Flash Point	Cool 4 deg C	PSEG01	H21	30604	12/17/2019	1010A
12/24/2019	Water	K6355-02 A	Flash Point	Cool 4 deg C	PSEG01	H21	13954	12/17/2019	1010A

Date/Time 12/18/19 10:30
 Received by: KS
 Relinquished by: DL

Date/Time 12/18/19 12:25
 Received by: DL
 Relinquished by: KS