



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106824

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/18/2019	13:05
2	CAL2	1	Water	NA	NA	20.1	7.00	12/18/2019	13:06
3	CAL3	1	Water	NA	NA	20.1	10.03	12/18/2019	13:07
4	ICV	1	Water	NA	NA	20.1	7.00	12/18/2019	13:09
5	CCV1	1	Water	NA	NA	20.2	2.02	12/18/2019	13:10
6	K6364-01	1	Water	NA	NA	18.2	5.39	12/18/2019	13:11
7	K6364-02	1	Water	NA	NA	19.2	5.23	12/18/2019	13:12
8	K6364-03	1	Water	NA	NA	19.3	5.39	12/18/2019	13:15
9	K6369-01	1	Water	NA	NA	16.8	6.01	12/18/2019	13:16
10	K6369-02	1	Water	NA	NA	17.4	6.57	12/18/2019	13:19
11	K6369-03	1	Water	NA	NA	18.6	6.78	12/18/2019	13:20
12	K6369-04	1	Water	NA	NA	18.3	6.69	12/18/2019	13:21
13	K6369-04DUP	1	Water	NA	NA	18.4	6.70	12/18/2019	13:22
14	CCV2	1	Water	NA	NA	20.1	12.02	12/18/2019	13:25

LB106824

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6364 WorkList ID : 134345 Department : Wet-Chemistry Date : 12-18-2019 12:50:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2019	Water	K6364-01 -A	pH	Cool 4 deg C	STER07	D42	CELLS-A	12/17/2019	SM 4500-H B
12/17/2019	Water	K6364-02 -A	pH	Cool 4 deg C	STER07	D42	CELLS-B/C	12/17/2019	SM 4500-H B
12/17/2019	Water	K6364-03 -A	pH	Cool 4 deg C	STER07	D42	CELLS-D	12/17/2019	SM 4500-H B
12/16/2019	Water	K6369-01 -C	pH	Cool 4 deg C	LOCK01	D51	MW-1	12/16/2019	SM 4500-H B
12/16/2019	Water	K6369-02 -C	pH	Cool 4 deg C	LOCK01	D51	MW-3	12/16/2019	SM 4500-H B
12/16/2019	Water	K6369-03 -C	pH	Cool 4 deg C	LOCK01	D51	MW-4	12/16/2019	SM 4500-H B
12/16/2019	Water	K6369-04 -C	pH	Cool 4 deg C	LOCK01	D51	MW-2	12/16/2019	SM 4500-H B

Date/Time 12.18.19 13:00

Received by: JA

Relinquished by: AM

Date/Time 12.18.19 14:10

Received by: AM

Relinquished by: JD