



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB106828

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP80644
Turbidity Calibration std, 0NTU	WP80643
Turbidity Calibration std, 20NTU	WP80639
Turbidity Calibration std, 40NTU	WP80638
Turbidity Calibration std, 1NTU	WP80642
Turbidity Calibration std, 10NTU	WP80640
Turbidity Calibration std, 5NTU	WP80641
10 NTU Standard 500 ml	W2489

Intercept: 0.1078

Slope: 0.9224

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.000	12/18/2019	15:19
2	CAL2	1	1	0.992	12/18/2019	15:20
3	CAL3	5	1	4.860	12/18/2019	15:21
4	CAL4	10	1	9.600	12/18/2019	15:35
5	CAL5	20	1	18.200	12/18/2019	15:40
6	CAL6	40	1	37.100	12/18/2019	15:45



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB106828

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	9.857	12/18/2019	15:52
2	ICB		1	0.180	0.078	12/18/2019	15:57
3	CCV1	10	1	9.460	10.139	12/18/2019	16:02
4	CCB1		1	0.181	0.079	12/18/2019	16:07
5	LB106828BL		1	0.190	0.089	12/18/2019	16:12
6	K6369-01		1	0.469	0.392	12/18/2019	16:20
7	K6369-01DUP		1	0.470	0.393	12/18/2019	16:23
8	K6369-02		1	0.534	0.462	12/18/2019	16:27
9	K6369-03		1	0.497	0.422	12/18/2019	16:30
10	K6369-04		1	7.150	7.635	12/18/2019	16:34
11	CCV2	10	1	9.500	10.182	12/18/2019	16:39
12	CCB2		1	0.192	0.091	12/18/2019	16:45

LB 106828

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBDITY121819 WorkList ID : 134350 Department : Wet-Chemistry Date : 12-18-2019 14:11:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2019	Water	K6369-01	Turbidity	Cool 4 deg C	LOCK01	D51	MW-1	12/16/2019	SM2130 B
12/18/2019	Water	K6369-02	Turbidity	Cool 4 deg C	LOCK01	D51	MW-3	12/16/2019	SM2130 B
12/18/2019	Water	K6369-03	Turbidity	Cool 4 deg C	LOCK01	D51	MW-4	12/16/2019	SM2130 B
12/18/2019	Water	K6369-04	Turbidity	Cool 4 deg C	LOCK01	D51	MW-2	12/16/2019	SM2130 B

Date/Time 12-18-19 14:30
Received by: AK
Relinquished by: 12

Date/Time 12-18-19 16:56
Received by: 12
Relinquished by: AK