



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106836

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/18/2019	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	12/18/2019	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	12/18/2019	16:42
4	ICV	1	Water	NA	NA	20.2	7.00	12/18/2019	16:44
5	CCV1	1	Water	NA	NA	20.1	2.02	12/18/2019	16:45
6	K6363-01	1	Water	NA	NA	15.8	8.01	12/18/2019	16:46
7	K6363-01DUP	1	Water	NA	NA	15.7	8.02	12/18/2019	16:47
8	CCV2	1	Water	NA	NA	20.1	12.02	12/18/2019	16:49

RB 106836

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6363 WorkList ID : 134361 Department : Wet-Chemistry Date : 12-18-2019 16:34:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2019	Water	K6363-01- ☒	pH	Cool 4 deg C	PSEG01	H23	T-2-CONTAINMENT-SUMP	12/18/2019	9040C

Date/Time 12-18-19 1635
Received by: J
Relinquished by: DL

Date/Time 12-18-19 17155
Received by: DL
Relinquished by: J4