


LB106846
INSTRUMENT ID CV1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	161	0.000	161				
0.2	0.200	0.169	-0.031	889	0.0 %	889				
2.5	2.500	2.462	-0.038	10289	0.0 %	10289				
5.0	5.000	5.236	0.236	21661	0.0 %	21661				
7.5	7.500	7.333	-0.167	30260	0.0 %	30260				
10.0	10.000	10.009	0.009	41230	0.0 %	41230				

Ib106846 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Dec 2019 10:05:43					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	161	-	0.0000	7470A	PPB	19 Dec 2019 10:11:18	S	Std
0.2 - 1	SD-2	889	-	0.2000	7470A	PPB	19 Dec 2019 10:13:26	S	Std
2.5 - 1	SD-5	10289	-	2.5000	7470A	PPB	19 Dec 2019 10:15:34	S	Std
5.0 - 1	SD-0	21661	-	5.0000	7470A	PPB	19 Dec 2019 10:17:42	S	Std
7.5 - 1	SD-5	30260	-	7.5000	7470A	PPB	19 Dec 2019 10:19:49	S	Std
10.0 - 1	SD-0	41230	-	10.0000	7470A	PPB	19 Dec 2019 10:24:56	S	Std
ICV36 - 1	ICV36	15535	3.7414 ✓	-	7470A	PPB	19 Dec 2019 10:27:27	U	SMPL
ICB36 - 1	ICB36	44	-0.0371 ✓	-	7470A	PPB	19 Dec 2019 10:29:33	U	SMPL
CCV47 - 1	CCV47	20416	4.9320 ✓	-	7470A	PPB	19 Dec 2019 10:31:39	U	SMPL
CCB47 - 1	CCB47	68	-0.0312 ✓	-	7470A	PPB	19 Dec 2019 10:33:46	U	SMPL
CRA - 1	CRA	869	0.1641 ✓	-	7470A	PPB	19 Dec 2019 10:35:52	U	SMPL
HighStd - 1	HighStd	40401	9.8067 ✓	-	7470A	PPB	19 Dec 2019 10:37:58	U	SMPL
ChkStd - 1	ChkStd	28755	6.9660 ✓	-	7470A	PPB	19 Dec 2019 10:40:05	U	SMPL
PB125595BL - 1	PBW	55	-0.0344 ✓	-	7470A	PPB	19 Dec 2019 10:42:12	U	SMPL
PB125595BS - 1	LCSW	16229	3.9107 ✓	-	7470A	PPB	19 Dec 2019 10:44:19	U	SMPL
K6328-04 - 1	OIL-SAMPLE-2-BG24	77	-0.0290 ✓	-	7470A	PPB	19 Dec 2019 10:46:26	U	SMPL
K6328-04DUP - 1	OIL-SAMPLE-2-BG24DUP	101	-0.0232 ✓	-	7470A	PPB	19 Dec 2019 10:48:34	U	SMPL
K6328-04MS - 1	OIL-SAMPLE-2-BG24MS	16305	3.9293 ✓	-	7470A	PPB	19 Dec 2019 10:50:42	U	SMPL
K6328-04MSD - 1	OIL-SAMPLE-2-BG24MSD	16363	3.9434 ✓	-	7470A	PPB	19 Dec 2019 10:52:50	U	SMPL
CCV48 - 1	CCV48	20461	4.9430 ✓	-	7470A	PPB	19 Dec 2019 10:54:57	U	SMPL
CCB48 - 1	CCB48	81	-0.0281 ✓	-	7470A	PPB	19 Dec 2019 10:57:03	U	SMPL
K6328-04LX5 - 1		30	-0.0405 ✓	-	7470A	PPB	19 Dec 2019 10:59:10	U	SMPL
PB125595TB - 1	PB125551TB	153	-0.0105 ✓	-	7470A	PPB	19 Dec 2019 11:01:16	U	SMPL
K6300-01 - 1	SSI-MW-4	81	-0.0281 ✓	-	7470A	PPB	19 Dec 2019 11:03:22	U	SMPL
K6300-01DUP - 1	SSI-MW-4DUP	92	-0.0254 ✓	-	7470A	PPB	19 Dec 2019 11:05:28	U	SMPL
K6300-02 - 1	K6300-01MS	16301	3.9283 ✓	-	7470A	PPB	19 Dec 2019 11:07:34	U	SMPL
K6300-03 - 1	K6300-01MSD	16176	3.8978 ✓	-	7470A	PPB	19 Dec 2019 11:09:41	U	SMPL
K6300-01LX5 - 1		4	-0.0468 ✓	-	7470A	PPB	19 Dec 2019 11:11:48	U	SMPL
CCV49 - 1	CCV49	20329	4.9108 ✓	-	7470A	PPB	19 Dec 2019 11:13:56	U	SMPL
CCB49 - 1	CCB49	118	-0.0190 ✓	-	7470A	PPB	19 Dec 2019 11:16:04	U	SMPL