



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106851

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/19/2019	09:20
2	K6363-02		1	9	100.00	>212.0	>212.0	12/19/2019	09:50
3	K6363-02DUP		1	9	100.00	>212.0	>212.0	12/19/2019	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB106851

WorkList Name : fp121919 WorkList ID : 134381 Department : Wet-Chemistry Date : 12-19-2019 08:53:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/24/2019	Water	K6363-02	Flash Point	Cool 4 deg C	PSEG01	H23	T-2-CONTAINMENT-SUMP	12/18/2019	1010A

Date/Time 12/19/19 08:55
 Received by: KS
 Relinquished by: JP

Date/Time 12/19/19 10:20
 Received by: JP
 Relinquished by: KS