



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106855

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/19/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.01	12/19/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	12/19/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	12/19/2019	17:14
5	CCV1	1	Water	NA	NA	20.1	2.01	12/19/2019	17:15
6	K6107-05	1	Solid	20.02	20	20.8	8.27	12/19/2019	17:16
7	K6117-09	1	Solid	20.03	20	19.9	8.26	12/19/2019	17:18
8	K6117-09DUP	1	Solid	20.04	20	19.8	8.27	12/19/2019	17:19
9	CCV2	1	Water	NA	NA	20.1	12.02	12/19/2019	17:20

RB106855

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K6379 WorkList ID : 134404 Department : Wet-Chemistry Date : 12-19-2019 16:42:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2019	Solid	K6107-05	Corrosivity	Cool 4 deg C	PSEG05	H23	HR-06-121819	12/18/2019	9045D
12/19/2019	Solid	K6117-09	Corrosivity	Cool 4 deg C	PSEG05		TR-04-121919	12/19/2019	9045D

Date/Time 12.19.19 1645
Received by: Jr
Relinquished by: ATM

Date/Time 12-19-19 18:00
Received by: ATM
Relinquished by: Jr