



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106856

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/19/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.01	12/19/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	12/19/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	12/19/2019	17:14
5	CCV1	1	Water	NA	NA	20.2	2.01	12/19/2019	17:25
6	K6379-01	1	Water	NA	NA	12.6	7.27	12/19/2019	17:26
7	K6379-01DUP	1	Water	NA	NA	12.7	7.28	12/19/2019	17:27
8	K6379-02	1	Water	NA	NA	15.3	7.91	12/19/2019	17:30
9	K6379-03	1	Water	NA	NA	15.5	7.76	12/19/2019	17:32
10	K6379-04	1	Water	NA	NA	16.8	7.06	12/19/2019	17:35
11	K6382-01	1	Water	NA	NA	2.0	5.78	12/19/2019	17:36
12	K6382-02	1	Water	NA	NA	3.9	7.01	12/19/2019	17:37
13	K6382-03	1	Water	NA	NA	5.2	6.92	12/19/2019	17:39
14	K6382-04	1	Water	NA	NA	2.3	7.35	12/19/2019	17:40
15	K6382-05	1	Water	NA	NA	4.0	7.10	12/19/2019	17:42
16	CCV2	1	Water	NA	NA	20.1	12.02	12/19/2019	17:44
17	K6382-06	1	Water	NA	NA	3.7	7.23	12/19/2019	17:45
18	K6382-07	1	Water	NA	NA	2.6	7.30	12/19/2019	17:46
19	K6382-08	1	Water	NA	NA	2.6	7.12	12/19/2019	17:48
20	K6382-09	1	Water	NA	NA	5.1	7.23	12/19/2019	17:49

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	K6382-10	1	Water	NA	NA	5.4	7.26	12/19/2019	17:50
22	K6382-11	1	Water	NA	NA	8.7	5.98	12/19/2019	18:00
23	K6382-12	1	Water	NA	NA	3.6	7.14	12/19/2019	18:02
24	K6382-13	1	Water	NA	NA	9.0	7.05	12/19/2019	18:05
25	K6382-14	1	Water	NA	NA	10.2	6.87	12/19/2019	18:06
26	CCV3	1	Water	NA	NA	20.1	2.02	12/19/2019	18:10

WORKLIST(Hardcopy Internal Chain)

PH 206856

WorkList Name: PH K6374 WorkList ID: 134405 Department: Wet-Chemistry Date: 12-19-2019 16:43:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2019	Water	K6379-01	pH	Cool 4 deg C	TULL02	D61	MH-F	12/19/2019	9040C
12/19/2019	Water	K6379-02	pH	Cool 4 deg C	TULL02	D61	MH-G	12/19/2019	9040C
12/19/2019	Water	K6379-03	pH	Cool 4 deg C	TULL02	D61	MH-20A	12/19/2019	9040C
12/19/2019	Water	K6379-04	pH	Cool 4 deg C	TULL02	D61	MH-BNMM-1	12/19/2019	9040C
12/19/2019	Water	K6382-01	pH	Cool 4 deg C	PSEG01	H23	30656	12/19/2019	9040C
12/19/2019	Water	K6382-02	pH	Cool 4 deg C	PSEG01	H23	30658	12/19/2019	9040C
12/19/2019	Water	K6382-03	pH	Cool 4 deg C	PSEG01	H23	30609	12/19/2019	9040C
12/19/2019	Water	K6382-04	pH	Cool 4 deg C	PSEG01	H23	30613	12/19/2019	9040C
12/19/2019	Water	K6382-05	pH	Cool 4 deg C	PSEG01	H23	30614	12/19/2019	9040C
12/19/2019	Water	K6382-06	pH	Cool 4 deg C	PSEG01	H23	30615	12/19/2019	9040C
12/19/2019	Water	K6382-07	pH	Cool 4 deg C	PSEG01	H23	30616	12/19/2019	9040C
12/19/2019	Water	K6382-08	pH	Cool 4 deg C	PSEG01	H23	30617	12/19/2019	9040C
12/19/2019	Water	K6382-09	pH	Cool 4 deg C	PSEG01	H23	30618	12/19/2019	9040C
12/19/2019	Water	K6382-10	pH	Cool 4 deg C	PSEG01	H23	30619	12/19/2019	9040C
12/19/2019	Water	K6382-11	pH	Cool 4 deg C	PSEG01	H23	30552	12/19/2019	9040C
12/19/2019	Water	K6382-12	pH	Cool 4 deg C	PSEG01	H23	30620	12/19/2019	9040C
12/19/2019	Water	K6382-13	pH	Cool 4 deg C	PSEG01	H23	30610	12/19/2019	9040C
12/19/2019	Water	K6382-14	pH	Cool 4 deg C	PSEG01	H23	30577	12/19/2019	9040C

Date/Time: 12-19-19 16:45
 Received by: Jn
 Relinquished by: Jn
 Date/Time: 12-19-19 18:00
 Received by: AM
 Relinquished by: Jn