


LB106865

INSTRUMENT ID CV1

ISM02.4



Linear

A= 0.0000e+000

B= 2.0762e-004 Slope

C= 5.3211e-003 Intercept

Rho= 0.9989004

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	V.D
0.00	0.000	0.008	0.008	13	0.000	13					
0.05	0.050										
0.20	0.200	0.234	0.034	1103	0.0 %	1103					17
2.50	2.500	2.550	0.050	12257	0.0 %	12257					2
5.00	5.000	4.729	-0.271	22750	0.0 %	22750					-5
7.50	7.500	7.797	0.297	37528	0.0 %	37528					4
10.0	10.000	9.882	-0.118	47568	0.0 %	47568					-1

6
12/20/19

INSTRUMENT ID CV1

LB106865

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	13	-	0	ISM02.4	PPB	12/20/2019 10:03	Std
0.2 S0.2	1103	-	0.2	ISM02.4	PPB	12/20/2019 10:05	Std
2.5 S2.5	12257	-	2.5	ISM02.4	PPB	12/20/2019 10:08	Std
5 S5.0	22750	-	5	ISM02.4	PPB	12/20/2019 10:10	Std
7.5 S7.5	37528	-	7.5	ISM02.4	PPB	12/20/2019 10:12	Std
10 S10.0	47568	-	10	ISM02.4	PPB	12/20/2019 10:14	Std
ICV	ICV	19323	4.0173	-	ISM02.4	PPB	12/20/2019 10:18 SMPL
ICB	ICB	253	0.0579	-	ISM02.4	PPB	12/20/2019 10:20 SMPL
CCV059	CCV059	22870	4.7537	-	ISM02.4	PPB	12/20/2019 10:22 SMPL
CCB059	CCB059	297	0.067	-	ISM02.4	PPB	12/20/2019 10:24 SMPL
PB125622E	PBS622	111	0.0284	-	ISM02.4	PPB	12/20/2019 10:27 SMPL
K6203-01	MC0BS0	7914	1.6485	-	ISM02.4	PPB	12/20/2019 10:29 SMPL
K6203-02	MC0BS1	1350	0.2856	-	ISM02.4	PPB	12/20/2019 10:31 SMPL
K6203-03	MC0BS2	873	0.1866	-	ISM02.4	PPB	12/20/2019 10:33 SMPL
K6203-04	MC0BS6	850	0.1818	-	ISM02.4	PPB	12/20/2019 10:36 SMPL
K6203-05	MC0BS8	2944	0.6166	-	ISM02.4	PPB	12/20/2019 10:38 SMPL
K6203-06	MC0BS9	1236	0.2619	-	ISM02.4	PPB	12/20/2019 10:40 SMPL
K6203-07	MC0BT0	706	0.1519	-	ISM02.4	PPB	12/20/2019 10:43 SMPL
K6203-08	MC0BT1	791	0.1696	-	ISM02.4	PPB	12/20/2019 10:45 SMPL
K6203-09	MC0BT2	799	0.1712	-	ISM02.4	PPB	12/20/2019 10:47 SMPL
K6203-10	MC0BT7	971	0.2069	-	ISM02.4	PPB	12/20/2019 10:49 SMPL
K6203-11	MC0BW1	958	0.2042	-	ISM02.4	PPB	12/20/2019 10:52 SMPL
K6203-12	MC0BW2	827	0.177	-	ISM02.4	PPB	12/20/2019 10:54 SMPL
K6203-13	MC0BW7	-3	0.0047	-	ISM02.4	PPB	12/20/2019 10:56 SMPL
K6203-14	MC0BW8	957	0.204	-	ISM02.4	PPB	12/20/2019 10:58 SMPL
K6203-15	MC0BW8D	947	0.2019	-	ISM02.4	PPB	12/20/2019 11:01 SMPL
K6203-16	MC0BW8S	13680	2.8456	-	ISM02.4	PPB	12/20/2019 11:04 SMPL
K6203-17	MC0BX0	5255	1.0964	-	ISM02.4	PPB	12/20/2019 11:06 SMPL
K6203-18	MC0BX1	1356	0.2869	-	ISM02.4	PPB	12/20/2019 11:09 SMPL
K6203-19	MC0BY9	936	0.1997	-	ISM02.4	PPB	12/20/2019 11:11 SMPL
CCV060	CCV060	23130	4.8077	-	ISM02.4	PPB	12/20/2019 11:13 SMPL
CCB060	CCB060	84	0.0228	-	ISM02.4	PPB	12/20/2019 11:15 SMPL
K6203-20	MC0BZ2	893	0.1907	-	ISM02.4	PPB	12/20/2019 11:18 SMPL
K6203-21	MC0BZ5	727	0.1563	-	ISM02.4	PPB	12/20/2019 11:20 SMPL
K6203-22	MC0BZ8	737	0.1583	-	ISM02.4	PPB	12/20/2019 11:22 SMPL
PB125623E	PBS623	90	0.024	-	ISM02.4	PPB	12/20/2019 11:25 SMPL
K6206-01	MC0BZ9	1519	0.3207	-	ISM02.4	PPB	12/20/2019 11:27 SMPL
K6206-02	MC0C27	880	0.188	-	ISM02.4	PPB	12/20/2019 11:29 SMPL
K6206-03	MC0C30	1151	0.2443	-	ISM02.4	PPB	12/20/2019 11:31 SMPL
K6206-04	MC0C31	1	0.0055	-	ISM02.4	PPB	12/20/2019 11:34 SMPL
K6206-05	MC0C34	1412	0.2985	-	ISM02.4	PPB	12/20/2019 11:36 SMPL
K6206-06	MC0C35	1409	0.2979	-	ISM02.4	PPB	12/20/2019 11:38 SMPL
K6206-07	MC0C40	3098	0.6485	-	ISM02.4	PPB	12/20/2019 11:40 SMPL
K6206-08	MC0C41	2827	0.5923	-	ISM02.4	PPB	12/20/2019 11:43 SMPL
K6206-09	MC0C46	3419	0.7152	-	ISM02.4	PPB	12/20/2019 11:45 SMPL
K6206-10	MC0C48	5047	1.0532	-	ISM02.4	PPB	12/20/2019 11:47 SMPL

INSTRUMENT ID CV1

LB106865

K6206-11	MC0C53	3426	0.7166 -	ISM02.4	PPB	12/20/2019 11:49	SMPL
K6206-12	MC0C54	1297	0.2746 -	ISM02.4	PPB	12/20/2019 11:52	SMPL
K6206-13	MC0C55	1814	0.382 -	ISM02.4	PPB	12/20/2019 11:54	SMPL
K6206-14	MC0C58	1718	0.362 -	ISM02.4	PPB	12/20/2019 11:56	SMPL
K6206-15	MC0C59	1791	0.3772 -	ISM02.4	PPB	12/20/2019 12:00	SMPL
K6206-16	MC0C59D	1766	0.372 -	ISM02.4	PPB	12/20/2019 12:05	SMPL
CCV061	CCV061	22903	4.7606 -	ISM02.4	PPB	12/20/2019 12:07	SMPL
CCB061	CCB061	-1	0.0051 -	ISM02.4	PPB	12/20/2019 12:09	SMPL
K6206-17	MC0C59S	14913	3.1016 -	ISM02.4	PPB	12/20/2019 12:11	SMPL
K6206-18	MC0C60	1458	0.308 -	ISM02.4	PPB	12/20/2019 12:15	SMPL
K6206-19	MC0C67	-5	0.0043 -	ISM02.4	PPB	12/20/2019 12:17	SMPL
K6206-20	MC0C68	1754	0.3695 -	ISM02.4	PPB	12/20/2019 12:19	SMPL
K6206-21	MC0C69	1314	0.2781 -	ISM02.4	PPB	12/20/2019 12:22	SMPL
K6206-22	MC0C70	1809	0.3809 -	ISM02.4	PPB	12/20/2019 12:24	SMPL
PB125624E	PBS624	11	0.0076 -	ISM02.4	PPB	12/20/2019 12:26	SMPL
K6301-01	MC0C71	0	0.0053 -	ISM02.4	PPB	12/20/2019 12:28	SMPL
K6301-02	MC0C72	633	0.1367 -	ISM02.4	PPB	12/20/2019 12:31	SMPL
K6301-03	MC0C80	2123	0.4461 -	ISM02.4	PPB	12/20/2019 12:33	SMPL
K6301-04	MC0C84	1447	0.3058 -	ISM02.4	PPB	12/20/2019 12:35	SMPL
K6301-05	MC0C85	1413	0.2987 -	ISM02.4	PPB	12/20/2019 12:37	SMPL
K6301-06	MC0C92	806	0.1727 -	ISM02.4	PPB	12/20/2019 12:40	SMPL
K6301-07	MC0C93	1030	0.2192 -	ISM02.4	PPB	12/20/2019 12:42	SMPL
K6301-08	MC0C93D	1038	0.2208 -	ISM02.4	PPB	12/20/2019 12:44	SMPL
K6301-09	MC0C93S	15646	3.2538 -	ISM02.4	PPB	12/20/2019 12:46	SMPL
K6301-10	MC0C94	3528	0.7378 -	ISM02.4	PPB	12/20/2019 12:49	SMPL
K6301-11	MC0C95	3729	0.7796 -	ISM02.4	PPB	12/20/2019 12:51	SMPL
CCV062	CCV062	22578	4.6931 -	ISM02.4	PPB	12/20/2019 12:53	SMPL
CCB062	CCB062	2	0.0057 -	ISM02.4	PPB	12/20/2019 12:56	SMPL