



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106870

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/20/2019	12:20
2	CAL2	1	Water	NA	NA	20.1	7.01	12/20/2019	12:21
3	CAL3	1	Water	NA	NA	20.1	10.03	12/20/2019	12:22
4	ICV	1	Water	NA	NA	20.1	6.99	12/20/2019	12:24
5	CCV1	1	Water	NA	NA	20.1	2.01	12/20/2019	12:25
6	K6386-01	1	Solid	20.02	20	18.1	8.25	12/20/2019	12:30
7	K6386-01DUP	1	Solid	20.03	20	18.2	8.26	12/20/2019	12:31
8	CCV2	1	Water	NA	NA	20.1	12.01	12/20/2019	12:35

13106870

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6386 WorkList ID : 134438 Department : Wet-Chemistry Date : 12-20-2019 12:11:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2019	Solid	K6386-01 -1	pH	Cool 4 deg C	TULL01	D61	STOCKPILE	12/19/2019	9045D

Date/Time 12-20-19 12:15
 Received by: Ja
 Relinquished by: Dk

Date/Time 12-20-19 13:15
 Received by: Dk
 Relinquished by: Ja