



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106878

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/20/2019	15:25
2	CAL2	1	Water	NA	NA	20.1	6.99	12/20/2019	15:26
3	CAL3	1	Water	NA	NA	20.1	10.03	12/20/2019	15:27
4	ICV	1	Water	NA	NA	20.1	7.00	12/20/2019	15:29
5	CCV1	1	Water	NA	NA	20.1	2.01	12/20/2019	15:30
6	K6393-01	1	Water	NA	NA	15.4	7.14	12/20/2019	15:32
7	K6393-01DUP	1	Water	NA	NA	15.5	7.15	12/20/2019	15:33
8	K6393-02	1	Water	NA	NA	15.8	8.01	12/20/2019	15:35
9	K6393-03	1	Water	NA	NA	16.5	7.69	12/20/2019	15:36
10	K6393-04	1	Water	NA	NA	17.2	7.95	12/20/2019	15:37
11	CCV2	1	Water	NA	NA	20.1	12.01	12/20/2019	15:39

LB 106876

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH K6393 WorkList ID: 134448 Department: Wet-Chemistry Date: 12-20-2019 15:17:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2019	Water	K6393-01 - E	pH	Cool 4 deg C	TULL02	D51	MH-BNM-1	12/20/2019	9040C
12/20/2019	Water	K6393-02 - E	pH	Cool 4 deg C	TULL02	D51	MH-F	12/20/2019	9040C
12/20/2019	Water	K6393-03 - E	pH	Cool 4 deg C	TULL02	D51	MH-20A	12/20/2019	9040C
12/20/2019	Water	K6393-04 - E	pH	Cool 4 deg C	TULL02	D51	MH-G	12/20/2019	9040C

Date/Time 12.20.19 15:20
Received by: JN
Relinquished by: DK

Date/Time 12.20.19 16:20
Received by: De
Relinquished by: Jh