



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106882

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2568         |
| buffer solution pH 7 yellow    | W2447         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 12/20/2019 | 16:30     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 12/20/2019 | 16:31     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 12/20/2019 | 16:32     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 6.99        | 12/20/2019 | 16:35     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.01        | 12/20/2019 | 16:36     |
| 6   | K6395-04    | 1  | Solid  | 20.03       | 20          | 19.2             | 8.50        | 12/20/2019 | 16:37     |
| 7   | K6395-04DUP | 1  | Solid  | 20.04       | 20          | 19.1             | 8.51        | 12/20/2019 | 16:38     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 12/20/2019 | 16:40     |