



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/24/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:33
In Date: 12/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:34
Out Date: 12/24/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106891

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6403-01	MH1561	1	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-02	MH1562	2	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-03	MH1563	3	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-04	MH1564	4	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-05	MH1565	5	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-06	MH1565D	6	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-07	MH1565S	7	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-08	MH1566	8	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-09	MH1567	9	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-10	MH1568	10	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-11	MH1569	11	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-12	MH1570	12	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-13	MH1571	13	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-14	MH1572	14	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-15	MH1573	15	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-16	MH1574	16	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS
K6403-17	MH1575	17	1.00	2.00	2.00	100.0	DUST SOLIDS 100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$