



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/24/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 12/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:37
Out Date: 12/24/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106892

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6408-01	MH0N65	1	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-02	MH13T0	2	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-03	MH13T0D	3	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-04	MH13T0S	4	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-05	MH14E0	5	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-06	MH14E1	6	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-07	MH14E2	7	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-08	MH14E3	8	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-09	MH14E4	9	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-10	MH14E5	10	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-11	MH14E6	11	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-12	MH14E7	12	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-13	MH14E9	13	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-14	MH14F0	14	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-15	MH14F1	15	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-16	MH14F2	16	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-17	MH14F3	17	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-18	MH14F4	18	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-19	MH14F5	19	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-20	MH14F6	20	1.00	2.00	2.00	100.0	100% SOLIDS
K6408-21	MH14F7	21	1.00	2.00	2.00	100.0	100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$