



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106896

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/23/2019	13:35
2	CAL2	1	Water	NA	NA	20.2	7.00	12/23/2019	13:36
3	CAL3	1	Water	NA	NA	20.2	10.04	12/23/2019	13:36
4	ICV	1	Water	NA	NA	20.1	7.01	12/23/2019	13:37
5	CCV1	1	Water	NA	NA	20.1	2.01	12/23/2019	13:38
6	K6399-01	1	Solid	20.03	20	21.1	6.33	12/23/2019	13:40
7	K6399-01DUP	1	Solid	20.04	20	21.2	6.34	12/23/2019	13:41
8	K6400-01	1	Solid	20.02	20	20.8	6.84	12/23/2019	13:42
9	CCV2	1	Water	NA	NA	20.1	12.02	12/23/2019	13:43

LB 106896

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6400 WorkList ID : 134500 Department : Wet-Chemistry Date : 12-23-2019 13:25:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2019	Solid	K6399-01 -C	pH	Cool 4 deg C	RESO01	D61	161-CHRUCH-ST-NORTH-HA	12/19/2019	9045D
12/19/2019	Solid	K6400-01 -C	pH	Cool 4 deg C	RESO01	D32	115-MAIDEN-LANE-BERGEN	12/19/2019	9045D

Date/Time 12.23.19 13:30
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 12.23.19 14:40
Received by: [Signature]
Relinquished by: [Signature]