



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106898

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/23/2019	13:35
2	CAL2	1	Water	NA	NA	20.2	7.00	12/23/2019	13:36
3	CAL3	1	Water	NA	NA	20.2	10.04	12/23/2019	13:37
4	ICV	1	Water	NA	NA	20.1	7.01	12/23/2019	13:37
5	CCV1	1	Water	NA	NA	20.2	2.01	12/23/2019	13:45
6	K6416-01	1	Water	NA	NA	18.4	4.64	12/23/2019	13:47
7	K6416-01DUP	1	Water	NA	NA	18.3	4.65	12/23/2019	13:48
8	CCV2	1	Water	NA	NA	20.1	12.02	12/23/2019	13:50

LB 106898

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K6416	WorkList ID :	134499	Department :	Wet-Chemistry	Date :	12-23-2019 13:24:12		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2019	Water	K6416-0+ 11-	pH	Cool 4 deg C	STER17		JS9600	12/16/2019	SM 4500-H B

Date/Time 12.23.19 13:25
Received by: JG
Relinquished by: ed

Date/Time 12.23.19 14:30
Received by: CO
Relinquished by: Jo