



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB106899

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: PhMeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP77570

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB106899BL		1	21.3	0.98	1.05	12/23/2019	13:50
2	LB106899BS	1413	1	20.8	1285.00	1397.07	12/23/2019	13:59
3	K6416-01		1	20.6	2.14	2.34	12/23/2019	14:10
4	K6416-01DUP		1	20.7	2.12	2.31	12/23/2019	14:17

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

UB106899

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CON122319 WorkList ID : 134501 Department : Wet-Chemistry Date : 12-23-2019 13:27:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/26/2019	Water	K6416-01	Conductance	Cool 4 deg C	STER17	JS9600		12/16/2019	SM2510 B

Date/Time 12-23-19 13:28
Received by: AW
Relinquished by: S

Date/Time 12-23-19 14:20
Received by: S
Relinquished by: AK